



Certificate of Analysis

Product Name:	Crystel Emerald	CRY 309
Reference:	CRY/EME/019	
Batch Number:	701902	
Date of Manufacture:	19-JAN-17	
Expiry Date:	JAN-2019	

This is to certify that Crystel Emerald solution above was tested on 19-JAN-17 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear pale brown solution	Conforms
SPECIFIC GRAVITY	1.020-1.040	1.034
pH (NEAT)	5.0-6.0	5.3
STERILITY	NO GROWTH	NA

Completed By:

Reviewed By:

Quality Department: E. Andrews

[Signature]

Date: 27 Jan 2017

Date: 27 JAN 17

DRF1574 CRY/EME/007/Issue3/August 2012

