

# Tristel Trio™

## Wipes System

COMPATIBILITY 

### TRISTEL TRIO WIPES SYSTEM MATERIAL COMPATIBILITY WITH ULTRASOUND MEDICAL DEVICES

Manufacturer	Instrument Type	Instrument Series	Model
Advanced Tactile Imaging	Ultrasound Probe	No Series Specified	Vaginal Tactile Imager (VTI)
BK Medical / BK Ultrasound	Ultrasound Probe	BK3000/BK5000	6C2, 6C2s, 9C2, 14L3, 13L4W, 10L2W, 18LS, E14C4t, E14C4, E14 CL4b, E10C4, 20R3, N13C5, 5Pl
		Flex Focus & Pro Focus	8814, 8815, 8816, 8824, 8836, 8862, 8863, 2050, 2052, 8808, 8808e, 8818, 8838, 8848
Carestream	Ultrasound Probe	No Series Specified	6C2, 6C2s, 9C2, 14L3, 13L4W, 10L2W, 18L5, E14C4t, E14C4, E14CL4b, E10C4, 20R3, N13C5, 5P1
Clarius Mobile Health	Handheld Ultrasound Scanners	No Series Specified	Clarius Scanner
			Arthrex Synergy MSK
Echosens	FibroScan and Ultrasound Unit	No Series Specified	M6.0, D5-1268, M3.5, D7-1257, M2.5, D10-1269
Esaote	Ultrasound Probe	No Series Specified	EC123*, EC1123*, IOE323*
	TOE Probe	No Series Specified	TEE022*, TEE122*
Exact Imaging	Ultrasound Probe	No Series Specified	EV9C, EV29L
Fujifilm	Ultrasound Probe	Ultrasonic Mini Probe System	PB-2020 M
GE Healthcare	Ultrasound Probe	Logiq Systems - E9	C1-5-D, C1-6-D, C2-9-D, M6C-D, IC5-9-D, 9L-D, 11L-D, ML6-15-D, L8-18i-D, 6s-D, 6s-D, RAB2-5-D, RAB4-8-D, RIC5-9-D
		Logiq Systems - S8	C1-5-D, IC5-9-D, 9L-D, ML6-15-D
		Logiq Systems - S7	C1-5-D, 8C, IC5-9-D, 9L-D, 11L-D, ML6-15, L8-18i-D, RAB4-8-D
		Logiq P Series	4C, 5CS, 8C, BE9CS, E8C, E8CS, 8L, 9L, 11L, ML6-15, I739, T739, 3S, 5S, 6S, 7S, 4D3C-L, 4D5C-L, 4DE7C
		Logiq C Series	C2-5-RC, E8CS-RC, L6-12-RC, 3S-RC
		Logiq A Series	3.5CS, 5CS, 8C, BE9CS, E8CS, 8L, 9L, 10L, 11L, 12L, I739, T739, 3S, 5S, 7S
		Logiq E Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, I12L-RS, L8-18i-RS, I739-RS, T739-RS, 3S-RS, 6S-RS
		Vivid E9	4C-D, 9L-D, 11L-D, ML6-15-D, 6S-D, 12S-D, 4VD
		Vivid P Series	4C, 5CS, 8C, E8CS, 8L, 11L, T739, 3S, 6S
		Vivid S Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, 3S-RS, 3Sc-RS, 5S-RS, 6S-RS, 7S-RS, 12S-RS, M4S-RS
		Vivid e/i/q Series	4C-RS, 8C-RS, E8C-RS, 8L-RS, 9L-RS, 12L-RS, 3S-RS, 3Sc-RS, 5S-RS, 6S-RS, 7S-RS, 12S-RS, M4S-RS
		Voluson Expert Series	AB2-7-D, 4C-D, C1-5-D, C4-8-D, M6C-D, IC5-9-D, 9L-D, 11L-D, ML6-15-D, 3S-D, RAB2-5-D, RAB4-8-D, RIC5-9-D, RIC6-12-D, RRE5-10-D
		Voluson Performance Series	4C-RS, E8C-RS, 12L-RS, 3Sc-RS, RAB2-5-RS, RIC5-9W-RS
		Voluson Signature Series	4C-RS, AB2-7-RS, C1-5-RS, 8C-RS, E8C-RS, 9L-RS, 12L-RS, ML6-15-RS, 3Sc-RS, RAB2-5-RS, RAB4-8-RS, RIC5-9W-RS
Venue 40 Series	4C-SC, E8CS-SC, 12L-SC, L8-18i-SC, 3S-SC		

Manufacturer	Instrument Type	Instrument Series	Model
GE Healthcare	TOE Probe	No Series Specified	TEE 6T, TEE 6T-RS, TEE 6Tc, TEE 6Tc-RS, TEE 6Tv TEE 9T, TEE 9T-RS, 6VT-D
Hitachi	Ultrasound Probe	No Series Specified	<p>C22l, C22K, C22P, C22T, C25, C251, C252, C25P, C35, C41, C41B, C41L47RP, C41RP, C41V, C41V1, C42, C42K, C42T, CC41R, CC41R1, CL4416R, L34, L43K, L44, L441, L44K, L44LA1, L46K, L46K1, L51K, L53K, L53K1, L55, L64, S12, S121, S21, S211, S22, S31, S31KP, S42, SML44, VC41V, ASU-1000B-3.5, ASU-1000C-3.5, ASU-1001, ASU-1002, ASU-1003, ASU-1005, ASU-1009, ASU-1010, ASU-1011, ASU-1012, ASU-1013, ASU-35-3, ASU-35CWD-2, ASU-35WL-10, ASU-35WSJ-7.5, ASU-36WL-10, ASU-64B, ASU-65B, ASU-67, UST-2265-2, UST-2266-5, UST-5024N-3.5, UST-5037P-3.5, UST-5039P-3.5, UST-5043P-3.5, UST-5044-3.5, UST-5045P-3.5, UST-5046-3.5, UST-5047-5, UST-52103, UST-52105, UST-52108, UST-52109, UST-52114P, UST-52124, UST-52127, UST-52129, UST-5226D-3.5/2, UST-5236D-2.5/2, UST-5242D-2.5/2, UST-5244-3.5, UST-5255-2.5, UST-5256-3.5, UST-5257-5, UST-5263D-2.5/2, UST-5265-2.5, UST-5266-3.5, UST-5267-5, UST-5268P-5, UST-5274-2.5, UST-5275-3.5, UST-5276-5, UST-5279-3.5, UST-5281-5, UST-5283-2.5, UST-5284-2.5, UST-5285-3.5, UST-5286-2.5, UST-5287-3.5, UST-5292-5, UST-5294-5, UST-5295, UST-5296, UST-5297, UST-5298, UST-5299, UST-533, UST-534, UST-535, UST-536, UST-5410, UST-5410H, UST-5411, UST-5412, UST-5413, UST-5415, UST-5417, UST-5419, UST-547, UST-5512U-7.5, UST-5514DTU-7.5, UST-5515-7.5, UST-5521L-7.5, UST-5523L-7.5, UST-5524-5, UST-5524-7.5, UST-5526L-7.5, UST-5531, UST-5534T-7.5, UST-5539-7.5, UST-5540P-7.5, UST-5542, UST-5543, UST-5545, UST-5546, UST-5548, UST-5551, UST-5561-7.5, UST-556T-7.5, UST-556TU-7.5, UST-567, UST-568, UST-5710-7.5, UST-5711-7.5, UST-5712, UST-5713T, UST-574T-7.5, UST-575-7.5, UST-576T-7.5, UST-577-7.5, UST-579T-7.5, UST-5818-5, UST-5819T-5, UST-657-5, UST-660-7.5, UST-665P-5, UST-667P-5, UST-670P-5, UST-675P, UST-676P, UST-677P, UST-9101-7.5, UST-9102-3.5, UST-9102U-3.5, UST-9103-5, UST-9104-5, UST-9111-5, UST-9112-5, UST-9113P-3.5, UST-9114-3.5, UST-9115-5, UST-9116P-5, UST-9118, UST-9119, UST-9120, UST-9121, UST-9123, UST-9124, UST-9125-7.5, UST-9126, UST-9127, UST-9128, UST-9130, UST-9132l, UST-9132T, UST-9133, UST-9135P, UST-9136, UST-9136U, UST-9137, UST-9145, UST-9146l, UST-9146T, UST-9147, UST-9148, UST-9149, UST-9151, UST-932-5, UST-932-7.5, UST-935N-5, UST-939-3.5, UST-939D-3.5, UST-943-3.5, UST-944B-3.5, UST-945B-5, UST-945BP-5, UST-952-5, UST-953P-5, UST-954-5, UST-957-7.5, UST-961-5, UST-961-7.5, UST-964P-5, UST-968P-5, UST-969-5, UST-970-7.5, UST-972-3.5, UST-974-5, UST-975-7.5, UST-977-5, UST-978-3.5, UST-979-3.5, UST-980-5, UST-980P-5, UST-981-5, UST-981P-5, UST-984-5, UST-984P-5, UST-985-7.5, UST-987-7.5, UST-988-2.5, UST-989-3.5, UST-989D-3.5, UST-990-5, UST-992-5, UST-994P-5, UST-995-7.5, UST-996P-5, UST-MC11-8731</p> <p>EUP-B314, EUP-B512, EUP-B514, EUP-B712, EUP-B715, EUP-C321T, EUP-C511, EUP-C514, EUP-C516, EUP-C524, EUP-C532, EUP-C715, EUP-CV714, EUP-F334, EUP-F531, EUP-L34T, EUP-L52, EUP-L53, EUP-L53L, EUP-L53S, EUP-L54MA, EUP-L73S, EUP-L74M, EUP-L75, EUP-LS85, EUP-O33JA, EUP-O53T, EUP-O54J, EUP-O732T, EUP-S50A, EUP-S52, EUP-S63, EUP-S70, EUP-S72, EUP-S80, EUP-V53W, EUP-V73W, EUP-VV731</p>

Manufacturer	Instrument Type	Instrument Series	Model
Medcom	Ultrasound Probe	Bi Plane Transrectal Transducer	BIPL7.5/70/128, BIC6.5/10/128
Mindray	Ultrasound Probe	No Series Specified	C5-1s, C5-1E, C5-1U, C5-2, C5-2s, C5-2E, C6-2E, C6-2Gs, C6-2GE, C7-3E, C11-3E, C11-3s, D7-2, D7-2E, D7-2s, D6-2, D6-2P, L7-3, L7-3s, L7-3E, L9-3U, L11-3U, L11-4, L11-4s, L12-3, L12-3E, L12-3s, L12-4, L12-4s, L14-6, L14-6s, L14-6P, L14-6N, L14-6Ns, L14-6NP, L14-6NE, L14-6WE, L14-6Ws, L14-6WU, LM14-6E, LM14-6s, LM16-4U, P4-2, P4-2s, P4-2E, P4-2NE, P10-4E, P10-4s, SC6-1U, SP5-1U, SP5-1s, SP5-1E, V11-3HU, V11-3U, 3C1, 3C1s, 3C1p, 10L24EA, 35C50EA, 35C50EB, 35C50EA, 35C50P, 65C15EAV, 65C15EA, 75L38EA, 75L38EB, 75L53EA, 75L60EA, 75L38P
	TOE Probe	No Series Specified	P7-3TE
Philips*	Ultrasound Probe	iU22	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD15	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD11 XE	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD9	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD7 XE	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD7	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD6	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		HD5	C8-4v (p/n 4535611 7292x where x is 3 or greater and p/n 4535612 8750x where x is 2 or greater)
		CX50	C9-3v
		Sparq	
		CX30	
		HD15	BP10-5ec
		HD11 XE	
		Affinity 50	
		Affinity 70	C9-4v
		EPIQ7	
		EPIQ5	
		Affinity 50	
		Affinity 70	
		CX50	C10-3v
		EPIQ7	
		EPIQ5	
		iU22	
		Affinity 70	
		CX50	S5-1
		EPIQ7	
		EPIQ5	
Affinity70			
CX50			
iU22			
iE33			

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Ultrasound Probe	EPIQ7	S12-4
		EPIQ5	
		Affinity70	
		Affinity50	
		iE33	
		HD15	
		HD11	
		CX50	
		HD11 XE	
		iE33/iU22	
		EPIQ7/iE33	X5-1
		EPIQ7/iE33/iU22	X7-2
		EPIQ7	C5-1
		QPIQ5	
		Affinity 70	
		iU22	
		iE33	
		HD15	
		iE33	C5-2
		iU22	
		HD15	
		HD11 XE	
		HD11	
		HD9	
		HD7 XE	
		HD7	
		HD6	
		HD5	
		iE33	C9-4v
		EPIQ7	
		EPIQ5	
		Affinity50	
		Affinity70	
CX50			
EPIQ7	C8-5		
EPIQ5			
Affinity 70			
Affinity 50			
iE33			
iU22			

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	Ultrasound Probe	HD7 XE	C8-5
		HD11 XE	
		HD9	
		HD11	
		HD15	
		CX30	
		CX50	
		iE33	L8-4
		iU22	
		HD11 XE	
		iE33	L9-3
		iU22	
		HD15	
		HD9	
		HD11 XE	L11-3
		iE33	
		EPIQ7	L15-7IO
		EPIQ5	
		Affinity 70	
		Affinity 50	
	iE33		
	iU22		
	HD11		
	HD15		
	CX30		
	CX50		
	TOE Probe	iE33	T6207
		iE33	D2cwc
		EPIQ7	
		EPIQ5	
		iU22	
		CX50	
		HD11	
HD9			
HD15			
HD11 XE			
No Series Specified		X8-2t, X8-2ti	
EPIQ7		D2tcd	
EPIQ5			
iE33			
iU22			
HD11 XE			

Manufacturer	Instrument Type	Instrument Series	Model
Philips*	TOE Probe	EPIQ7	S8-3
		EPIQ5	
		Affinity50	
		Affinity70	
		CX30	
		CX50	
		iE33	
		HD11	
		HD11 XE	
		HD15	
		iE33	S7-2t
		iU22	
		HD15	
		HD11 XE	
		HD11	
		HD9	
		HD7 XE	
		HD7	
		HD6	
		HD5	
		iE33	S7-2 Omni (21778A)
		iU22	
		HD15	
		HD11 XE	
		HD11	
		HD9	
		HD7 XE	
		HD7	
		HD6	
		HD5	
		iE33	T6H/21378A OMNI III
		HD11 XE	
		HD7 XE	
EPIQ7	X7-2t		
EPIQ5			
Affinity70			
iE33			
CX50			
Sparq			

Manufacturer	Instrument Type	Instrument Series	Model
Samsung*	Ultrasound Probe	No Series Specified	ER4-9/10ED, EV4-9/10ED, V5-9, 3D4-9(ES), EVN4-9/10ED, CA1-7A, C2-6(IC), CA2-8A, CF2-8, CA2-8AD, L5-13(IS), L7-16(IS), LA3-16AI, L3-12A, LA3-14AD, L5-13/50, CV1-8A, CV1-8AD, 3DC-6(EN), 3D4-9(ES), EV3-10B, C2-8(ED), CA2-9A, CA2-9AD, CA1-7AD, CA2-6BM, EVN4-9, EC4-9(CD), L3-8(ED), LS6-15, LA3-14AD, LA3-14A, PA1-5A, PA1-5AD, SP3-8, CA3-10A, VR5-9, E3-12A, EA2-11B
Siemens*	Ultrasound Probe	Acuson Antares	CH4-1, CH6-2, CW2, EC9-4, PH4-1
		Acuson CV70	3V2c, 7V3c, CW2, CW5
		Acuson CypresS	3V2c, 4C1, 7V3c, Aux CW
		Acuson S1000	4C1, 4V1, 6C2, 8V3, CW2, CW5, EC9-4, EV-8C4
		Acuson S2000	4C1, 4V1, 4V1c, 6C2, 8V3, CW2, CW5, EC9-4, EV-8C4
		Acuson SC2000 Prime	4V1c, 8V3, AUX CW, CW2
		Acuson S3000	4C1, 4V1, 4V1c, 6C2, 8V3, CW2, CW5, EC9-4, EV-8C4
		Acuson SG2000	4V1c, 8V3, CW2
		Acuson Sequoia	15L8, 3V2C, 4C1, 4V1/4V1c, 5V2c, 6C2, 6C3, 7V3c, 8C4, 8L5, 8L5T, 8V3c, 8V5, Aux CW, EC-10C5, EV-8C4
		Acuson X700	4C1, 4C1v, 6C2, CW2, CW5
		Acuson X150	CH5-2, EC9-4, EV9-4, P4-2, P8-4
		Acuson X300	CH5-2, CW2, CW5, EC9-4, EV9-4, P4-2, P8-4
		Acuson X300PE	CH5-2, CW2, CW5, EC9-4, EV9-4, P4-2, P8-4
		Acuson X500	CW2, CW5, EC9-4
		Sonoline G20	EC9-4
	Sonoline G40	CH5-2, EC9-4, EV9-4, P4-2	
	Sonoline G50/G60	CW2, CW5, EC9-4	
	TOE Probe	Acuson Cypress	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S1000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S2000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson S3000	V5Ms (SN 327xxxxx & 328xxxxx)
		Acuson SC2000	V5Ms (SN 327xxxxx & 328xxxxx)
Acuson SC2000 Prime		V5Ms (-SN 327XXXXX)	
Acuson SC2000 Prime		V5Ms (SN 328XXXXX -)	
Acuson Sequoia		V5Ms (SN 327xxxxx & 328xxxxx)	
Acuson X300PE		V5Ms (SN 327xxxxx & 328xxxxx)	
Acuson X700		V5Ms (SN 327xxxxx & 328xxxxx)	
Acuson SC2000	Z6Ms		

Manufacturer	Instrument Type	Instrument Series	Model
Sonosite Fujifilm	Portable Ultrasound System	FC1-System	Powerplug
		M-Turbo-System	M-Turbo-System
			M-Turbo-TTC
			Foot Paddle
			Powerplug
			Mini-Dock
		NanoMaxx-System	NanoMaxx-System
			Dock
			Powerplug
		M Turbo M-OB/GYN Office	(System & Dock)
	No Series Specified	Micro Maxx	
	No Series Specified	TITAN	
	Skin surface Transducers	No Series Specified	SLT
	Skin surface Transducers in vascula	No Series Specified	P11x
			L52/L52e - Vet Only
	Ultrasound Probe	Edge-System	P11x2
		Edge II-System	P11x2
		S-System	P11x2
		SII-System	P11x2
		M-Turbo-System	P11x2
		No Series Specified	D2/D2x
	Ultrasound System	iViz System	iViz System
			Protective Cover
	Ultrasound Unit	Edge-System	Edge-TTC
			Powerplug
			Mini-Dock
		Edge II-System	Edge II-System
			Edge-TTC
			Powerplug
			Mini-Dock
		S-System	S-System
			S-Series-TTC
Powerplug			
SII-System		SII-System	
		Powerplug	
Ultrasound Unit System	X-Porte - System	Powerplug	
TOE Probe	No Series Specified	TEExi Transducer	



Manufacturer	Instrument Type	Instrument Series	Model
Toshiba	Ultrasound Probe	Aplio & Viamo Series	PC-20M, PC-50M, PET-508MA, PET-512MA/512MC, PET511BTM, PST-25BT/50AT/65AT, PVT-661VT, PVT-674BT, PVT-781VT, PVT-781VTE, PVL-715RST, PVT-770RT, PVT-745BTF, PVT-745BTH, PLT-308P, PLT-705BTF, PLT-705BTH, PVT-745BTV, PVT375ST *, PVT-745BTF/745BTH, PLT-308P, PVT-770RT, PLT-705BTF/705BTH, PVT-745BTV, PVT-7350BTP, PLT-1204BX, PLT-704SBT/705BT/1005BT/PVT-781VT, PLT704ST, PLT1204ST, PVT375ST, PST-25SX, PST25ST/50BT, PVT712BT, PST-25BT, PST-30BT, PST-65AT, PLT1204ST, PLT-1204MV, PVT-375MV/382MV/675MV/675MVL, PVT-681MV, PST25ST, PST50BT, PST-30BT, PVT-375BT, PVT-375SC, PVT-382BT, PVT712BT, PLT-604AT, PLT-704AT, PLT-8055AT, PLT-1202S, PLT-1204BT, PLT-1204BX, PLT-704SBT, PLT-705BT, PLT-1005BT, PLT704ST
		PVF-738F/738H/745V	
		PLG-506M/805S	
		PVG-366M/381M/681S/720S, PVG-600S/601V/630V, PLQ-703A/805A/1203A, PVQ-375A/381A/641V/662A/674A	
		PLM-503AT/703AT/805AT	
		PLM-1202S/1204AT	
		PVF-381MT/620ST/575AT	
		PVM-375AR/381AT/620ST/621VT	
		PVM-651VT/656AT/662AT/672AT	
		PSM-20CT/25AT/30BT	
		PVG-366M*, PVG-381M*, PVG-681S*, PVG-600S*, PVG-601V*, PVG-630V*, PVM-651VT, PVM-656A, PVM-740RT, PVL-715R, PVL-715RT, PVF-738F*, PVF-738H*, PVF-745V*, PLF-308P, PLG-308P, PLM-1202S, PLM-1204AT, PVF-381MT, PSM-37AT, PSM-70AT, PVF-575AT, PVQ-375A, PVQ-641V, PVQ-674A, PVF-3620ST PVM-375AT, PSM-37CT, PSM-50AT, PLG-506M*, PLG-805S*, PLQ-703A, PLQ-805A, PLQ-1203A, PLM-503AT, PLM-703AT, PLM-805AT, PSM-20CT, PSM-30BT	
	Xario Series	PSU-25BT/30BT & PVU-375BT/382BT	
PSU-50BT & PVU712BT			
PLU-704BT			
PVU-745BTV			
PVU-745BTF/745BTH/770ST & PVL-715RS			
PVU-674BT/781VT & PLU-1005BT			
PLU-1204BT			
PVL-715RS, PVU-745BTV, PVU-745BTF, PVU-745BTH, PVU-770ST, PVU-681MVL, PVU-375BT, PVU-382BT, PSU-50BT, PSU-25BT, PSU-30BT, PSU-70BT, PVU712BT, PLU-704BT			
TOE Probe	Aplio & Viamo Series	PET-508MA, PET-512MA, PET-512MC, PET511BTM	

\* Compatible with the Tristel Sporidical Wipe, part of the Tristel Trio Wipe System

**Disclaimer:** Material compatibility testing performed via different methods including wiping and soaking. Tested materials subjected to varying number of cycles.

For more information contact Tristel team on [cooperation@tristel.com](mailto:cooperation@tristel.com)

**Tristel**

**Created by:** Tristel Solutions Limited, Lynx Business Park, Cambs, UK, CB8 7NY  
T +44 (0) 1638 721500 - E [mail@tristel.com](mailto:mail@tristel.com) - W [www.tristel.com](http://www.tristel.com)

For Tristel patent information please visit: <http://www.our-patents.info/tristel>

**Australian sponsor:** Tristel Pty Ltd, 40/328 Reserve Road, Cheltenham, VIC 3192  
T 1300 680 898 - F +61 (0)3 9533 6193 - E [mail-au@tristel.com](mailto:mail-au@tristel.com)

**New Zealand:** Tristel New Zealand Limited, P.O Box 15451, Tauranga, New Zealand  
T +64 (0)7 5771560 - F +64 (0)7 5771567 - E [info@tristel.co.nz](mailto:info@tristel.co.nz)

**Hong Kong & Taiwan:** Tristel Asia Limited, 21st Floor, 168 Electric Road, North Point, Hong Kong  
T +852 2895 6968 - F +852 2869 4388

Copyright © Tristel Solutions  
Mkt-Val-449-2  
March 2018