







# **Certificate of Analysis**

Product Name:	Crystel Platinum CRY109
Reference:	CRY/PLA/019
Batch Number:	P624400
Date of Manufacture:	02-AUG-16
Expiry Date:	AUG-2018

This is to certify that Crystel Platinum solution above was tested on 04-AUG-16 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear solution free of matter	Conforms
SPECIFIC GRAVITY	0.900 - 1.100	0.990
pH (neat)	9.0 - 10.0	9.8
STERILITY	No growth	Conforms

Completed By:	Reviewed By:
Quality Department: Concleus	TENSC W
Date: 28 Soot 2016	Date:28 . Sep. (6

DRF1399 CRY/PLA/007/Issue3/June 2012





### **Sterility Testing Analytical Report**

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY109 Platinum

Lot No: P624400 ICN: 0916001D

Subcontractor's Report Number: 2016213032

Lab Book Reference: ST03-29 Date Received: 05 Sep 2016 Date Completed: 23 Sep 2016

No. of samples: 10

Sterility Test - Membrane Filtration

Stermey rest - Membrane Philadion		
Medium	Result	
Tryptone Soy Broth	No growth	
Fluid Thioglycollate Medium	. No growth	

### 

#### Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1			23 Sep 2016

CONFIDENTIAL

Page 1 of 1

ST03-29 REVISION NO: 1

ICN: 0916001D DATE: 23 Sep 2016

TSCTS044-1 Rev 2

UR
Deacon Park
Meorgate Road
Knowsley
Liverpool L33 7RN
Tel +44 (0)151 547 7444
Eav + 44 (0)151 547 7100

USA 124 Bernard E. Saint Jean Drive East Falmouth MA 02536-4445 USA Tel: 001 508 540 3444 Fax: 001 508 540 8680

Germany PYROQUANT DIAGNOSTIK GmbH Opelstrasse 14 D-64546 Mortelden-Walldorf Germany Tel: 0049 6105 96 10 0 Ex: 0019 6105 96 10 15



http://www.synergyhealthplc.com

### **Certificate of Irradiation**

Date Issued: 09-Sep-2015 UK33S11455485-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1075628	
Customer Reference Number:	700363	
Product Description:	BOTTLE 100ML DV4224 25-45kGy	
Validation Reference:	4224	
Quantity Received:	6	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	1pit	
lı .	rradiation Data	
Date and Time of Irradiation:	09-Sep-2015 03:37	
Reference Dose Range kGy:	32.3 - 32.3	
Calculated Minimum Dose kGy:	30.8	
Calculated Maximum Dose kGy:	35.2	

Irradiation Release Authorised By Synergy Health plc



http://www.synergyhealthplc.com

### **Certificate of Irradiation**

Date issued: 11-Aug-2016 UK33S11678232-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1075625	
Customer Reference Number:	702828	
Product Description:	CAP 001-G DV1.4225 25-45kGy	
Validation Reference:	1.4225	
Quantity Received:	3	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	1 pit	
	rradiation Data	
Date and Time of Irradiation:	11-Aug-2016 02:38	
Reference Dose Range kGy:	33.4 - 33.4	
Calculated Minimum Dose kGy:	29.0	
Calculated Maximum Dose kGy:	37.5	

Irradiation Release Authorised By Synergy Health plc



http://www.synergyhealthplc.com

## **Certificate of Irradiation**

Date Issued: 03-Nov-2015

UK33S11373223-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1094470	
Customer Reference Number:	0000015258	
Product Description:	BAG003-G VALIDATION	
Quantity Received:	5	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	1pit, VALIDATION	
Other Process Details:	This validation batch has received an underdose	
lı .	rradiation Data	
ate and Time of Irradiation: 03-Nov-2015 13:12		

Irradiation Release Authorised By Synergy Health plc