

CERTIFICATE OF ANALYSIS		
PRODUCT NAME:	Perastel Multishot (7 Day) Working Solution	
REFERENCE: MED/PER/042		
THIS IS TO CERTIFY THAT	Perastel Multishot (7 Day) Working Solution	
BEARING THE BATCH NUMBER P628601		
WAS TESTED ON	07-OCT-16	
AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS		
PROPERTIES	Specification	Results
Appearance	Pink liquid free from haziness or suspended matter	Conforms
Specific Gravity	1.010 +/- 0.010	1.011
На	1.0 – 3.0	1.5
Peracetic acid Content	> 400 ppm	451 ppm
Completed By:	Reviewed By:	
Quality Department: Carakeus		
Date: 17 Oct 2016 Date: 17- Oct - 2016		

Document No: QCA -MED/PER/042-5

DRF1926



CERTIFICATE OF ANALYSIS PRODUCT NAME: Perastel Multishot (7 Day) Base Solution REFERENCE: MED/PER/038 DATE OF MANUFACTURE: 16-SEP-16 **EXPIRY DATE:** MAR-2018 THIS IS TO CERTIFY THAT Perastel Multishot (7 Day) Base Solution P628601 BEARING THE BATCH NUMBER 16-SEP-16 WAS TESTED ON AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS **PROPERTIES** Specification Results Conforms **APPEARANCE** Clear, colourless liquid free of matter SPECIFIC GRAVITY 1.010 +/- 0.010 1.018 pΗ 1.0 - 2.01.5 Completed By: Reviewed By: Quality Department: Gandrews Date: 17 OCE 2016 Date: 17-OCT-2016

Document No: QCA-MED/PER/038-3 Page 1 of 1





CERTIFICATE OF ANALYSIS PRODUCT NAME: Perastel Multishot (7 Day) Activator Solution EX2 MED/PER/041 **REFERENCE:** DATE OF MANUFACTURE: 21-SEP-16 **EXPIRY DATE: MAR-2018** Perastel Multishot (7 Day) Activator Solution EX2 THIS IS TO CERTIFY THAT P628601 BEARING THE BATCH NUMBER 21-SEP-16 WAS TESTED ON AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS Results Specification **PROPERTIES** Conforms Pale pink liquid free of Appearance suspended matter Conforms Mild Lemon Like Odour 1.000 +/- 0.005 **Specific Gravity** 1.001 6.0 pН 5.0 - 7.5 Reviewed By: Completed By: Quality Department: ... Can Department Date: 17 Oct 2016 Date: 17-0cT-2016

Document No: QCA-MED/PER/041-4

DRF1512