



Certificate of Analysis

Product Name:	Crystel Silver CRY 104
Reference:	CRY/SIL/019
Batch Number:	P630700
Date of Manufacture:	31-OCT-16
Expiry Date:	OCT-2018

This is to certify that Crystel Silver solution above was tested on 31-OCT-16 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear and free from suspended matter	Conforms
ODOUR	Alcohol Like	Conforms
SPECIFIC GRAVITY	0.875 – 0.895	0.880
STERILITY	No growth	Conforms
IRRADIATION	Achieves minimum dose	Conforms
ENDOTOXIN	<0.25EU/ml	<0.1
	Recovery 50-200%	117%

Completed By:

Quality Department: E Andrews

Date: 19 Dec 2016

Reviewed By:

[Signature]

Date: 19/12/16

DRF1574 CRY/SIL/007/Issue5/Sept 2012





Specialists in Endotoxin and Glucan Detection

Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY104 Silver	Lab Book Reference: ST03-45
Lot No: P630700	Date Received: 22 Nov 2016
ICN: 1116031F	Date Completed: 13 Dec 2016
Subcontractor's Report Number: 2016269048	No. of samples: 10

Sterility Test – Membrane Filtration

Medium	Result
Tryptone Soy Broth	No growth
Fluid Thioglycollate Medium	No growth

Conclusions:

The samples submitted comply with the test for sterility in accordance with the client's methodology.

Results Reported by:

[Signature]

 Laboratory Technician: Sophie Hughes

Date: *13 Dec 2016*

Reviewed by:

[Signature]

 Quality Assurance Manager
 UK General Manager
 Technical Manager
 Laboratory Manager

Date: *13 Dec 2016*

QA Approved by:

[Signature]

 Quality Assurance Manager
 UK General Manager
 Technical Manager

Date: *13 Dec 2016*

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	S. Hughes	13 Dec 2016

CONFIDENTIAL

Page 1 of 1

ST03-45
REVISION: NO: 1

ICN: 1116031F
DATE: 13 Dec 2016

TSCTS044-1 Rev 2

UK
 Deacon Park
 Moorgate Road
 Knowsley
 Liverpool L33 7RX
 Tel: +44 (0)151 547 7444
 Fax: +44 (0)151 547 7400

USA
 124 Bernard E. Saint Jean Drive
 East Falmouth,
 MA 02536-4445
 USA
 Tel: 001 508 540 3444
 Fax: 001 508 540 8680

Germany
 PYROQUANT DIAGNOSTIK GmbH
 Opelstrasse 14
 D-64546 Morfelden-Walldorf
 Germany
 Tel: 0049 6105 96 10 0
 Fax: 0049 6105 96 10 15

Web: www.acciiuk.co.uk

Registered in England and Wales, Company Registration Number: BR002906



Specialists in Endotoxin and Glucan Detection

**LIMULUS AMEBOCYTE LYSATE (LAL) ANALYTICAL REPORT
Release Test - Kinetic Turbidimetric Method (SOPCTS012)**

Sample ID: CRY104 SILVER
Lot No: P630700
ICN No: 1116031F

Lab Book Reference: EL187-54 (A)
Date Received: 22 NOV 2016
Date Completed: 25 NOV 2016

Reagents

Pyrotell®-T: Lot Number 516-05-784-T. Pyrotell®-T was reconstituted with Pyrosol® Buffer and used to Assay all dilutions.

Pyrosol® Buffer: Lot Number 226-391.

LAL Reagent Water (LRW): Lot Number AAH207245. LRW was used to reconstitute endotoxin, prepare dilutions of endotoxin and sample(s), and serve as the negative control.

Instrumentation: Pyros® Kinetix Incubating Kinetic Tube Reader; S/N; 3G-1166.

Control Standard Endotoxin (CSE): Lot number 152, *Escherichia coli* O113:H10. The potency of the CSE is 20 EU/ng when measured against Reference Standard Endotoxin (RSE), lot number H0K354 (Food and Drug Administration, CBER) with Pyrotell®-T lot number 516-04-784-T.

The sensitivity of the assay is the lowest concentration of CSE used to construct the calibration curve. A series of 10-fold dilutions between 0.001 and 1.0 EU/mL was tested in duplicate and the results were used to construct the standard curve.

Preparation and Testing of Sample: The samples were diluted with LRW and analysed at a 1:100 dilution, in accordance with validation EL168-06. Positive product controls were tested, in parallel, on sample dilutions spiked with an Endotoxin concentration of 0.1 EU/ml. The sample to lysate ratio was 4:1.

Results: The Release Test is performed according to a validated method and is used to release finished products. The test can also be performed for release of raw/in-process materials. The absolute value for the correlation coefficient for the regression of log onset time versus log endotoxin concentration was greater than or equal to 0.980. The negative control (LRW) was tested in duplicate and contained less than 0.001 EU/mL. Inhibition or enhancement of the bacterial Endotoxin test by the sample is absent if the concentration of the Endotoxin assayed in the Positive Product Controls are within 50% to 200% of the nominal concentration.

CONFIDENTIAL

Page 1 of 2

**EL187-54 (A)
REVISION NO: 1**

**ICN: 1116031F
DATE: 25 NOV 2016**

TSCTS015-3.1 Rev 17

UK
Deacon Park
Moorgate Road
Knowsley
Liverpool L33 7RN
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA
124 Bernard E. Saint Jean Drive
East Falmouth
MA 02536-4445
USA
Tel: 001 508 540 3444
Fax: 001 508 540 8680

Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Morleiden-Walldorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15

Web: www.acciuk.co.uk

Registered in England and Wales. Company Registration Number: BR002906



Sample Results:

Sample Identification	% Spike Recovery	EU/ml
CRY104 SILVER P630700	117	<0.1

All samples analysed passed the specification of <0.25 EU/ml.

Approval:

Technician:.....
Laboratory Technician: *Charlotte Ray*

Date: *25 NOV 2016*

Reviewed By: *D Solauer*

Date: *02 DEC 2016*

- Quality Assurance Manager
- UK General Manager
- Technical Manager
- Laboratory Manager

QA Approval:.....

Date: *02 DEC 2016*

- Quality Assurance Manager
- UK General Manager
- Technical Manager

References:

Associates of Cape Cod SOPCTS012, *General Kinetic LAL Turbidimetric Assay*
United States Pharmacopoeia <85> *Bacterial Endotoxins Test*
European Pharmacopoeia, *Section 2.6.14.*

Revision History:

DCCF No:	Revision:	Description:	Originator	Effective Date:
N/A	1	Revision 1: No Changes	C RAY	25 NOV 2016



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 13-Nov-2016

UK33S11747917-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products
 EN ISO 9001:2008 Quality Management System
 EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
 Unit 4C Lynx Bus Park, Fordham Rd
 Newmarket
 Suffolk CB8 7NY
 UNITED KINGDOM

Order Information

Account Number:	100708
Synergy Health Sales Part Reference:	1108717
Customer Reference Number:	703526
Product Description:	ALCOHOL 1 LTR TRIGGER SPRAY DV4778 20-45kGy
Validation Reference:	4778
Quantity Received:	140
Customer Minimum Specification kGy:	20.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	P630700, 3 Pits

Irradiation Data

Date and Time of Irradiation:	12-Nov-2016 21:46
Reference Dose Range kGy:	33.9 - 34.5
Calculated Minimum Dose kGy:	27.5
Calculated Maximum Dose kGy:	39.8

Irradiation Release Authorised By Synergy Health plc

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
 Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069