







Certificate of Analysis-TITANIUM WIPES-CRY/TTM/049

PRODUCT batch number: P700300

Product Name:	TITANIUM WIPE	TITANIUM WIPES Working Solution	
Reference:	CRY/TTM/047		
	IM WIPES Working Solution b -DEC-16 and found to comply and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS	
APPEARANCE	Clear pale yellow liquid free of matter	Conforms	
ODOUR	Chlorine like	Conforms	
ClO2 Concentration (ppm)	Conforms to QCF CRY/TTM/050	Conforms	

Product Name:		TITANIUM WIPES Base Solution	
Reference:		CRY/TTM/045	
Date of Manufacture:		08-DEC-16	
Date of Expiry:		DEC-2018	
This is to certify that TI DEC-16 and found to co and method of analysi	omply with all requ		o was tested on o8- ut in the specifications
PROPERTIES	SPECIFICATION		RESULTS
APPEARANCE	Clear liquid free of matter		Conforms
SPECIFIC GRAVITY	1.000-1.010		1.006
рН	2.0	- 3.0	2,2
STERILITY	No growth		Conforms

DRF2550 CRY/TTM/049/Issue1











Product Name:	TITANIUM WIPES Activator Solution	
Reference:	CRY/TTM/046	
Date of Manufacture:	01-DEC-16	
Date of Expiry:	DEC-2018	

This is to certify that TITANIUM WIPES Activator Solution 633602 was tested on 01-DEC-16 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear and free of matter	Conforms
pH	11.0-11.5	11.1
SODIUM CHLORITE	0.2% +/- 0.005%	0.197%
STERILITY	No growth	Conforms

Completed By:

Date: 27 Feb 2017

Reviewed By:

Date: 27 FES 17





Certificate of Analysis

Sample Details Product Name: Titanium burstable wipes

Batch Number: P700300

| Job Number: 15990 | Batch Code | P700300 | P

Sample ID:

56898

Registered date:

10-Feb-17

Completion date:

n date: 27-Feb-17

Data

Test Method	Test	Specification	Result	Units	Pass/Fail
Direct Inoculation Sterility Test (cFTM)	DI Sterility (cFTM)	Complies / Does Not Cornply	Complies	N/A	Pass
Direct Inoculation Sterility Test (TSB)	DI Sterility (TSB)	Complies / Does Not Comply	Complies	N/A	Pass
Membrane Filtration Sterility Test (cFTM)	MF Sterility (cFTM)	Complies / Does Not Cornply	Complies	N/A	Pass
Membrane Filtration Sterility Test (TSB)	MF Sterility (TSB)	Complies / Does Not Cornply	Complies	N/A	Pass

The product meets the following specification:

Sterility Test current Ph. Eur. by Membrane Filtration and Direct Inoculation. Validation report HVAL_R405.v1

LabHQ Product Specification Version:

Approval

Approved By Katie Pearson Approved Date: 27-Feb-2017 12:40:29

User ID kpearson

Title Laboratory Supervisor

This approval is performed by electronic signature in LabHQ LIMS and is the legal equivalent of the approver's hand-written signature.



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 28-Jan-2017 UK31S11786836-1-1

This is to certify that Elgin Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1103783	
Customer Reference Number:	703920	
Product Description:	BURST/TITANIUM WIPES 10-20KGY	
Validation Reference:	1.6922	
Quantity Received:	40	
Customer Minimum Specification kGy:	10.0	
Customer Maximum Specification kGy:	20.0	
Customer Unit Lot/Batch Number:	LOT NO. P700300, 1 PALLET	
li li	rradiation Data	
Date and Time of Irradiation:	28-Jan-2017 00:55	
Reference Dose Range kGy:	14.4 - 14.4	
Calculated Minimum Dose kGy:	11.7	
Calculated Maximum Dose kGy:	14.5	

Irradiation Release Authorised By Synergy Health plc