







Certificate of Analysis

Product Name:	Crystel Platinum RTU Solution CRY110	
Reference:	CRY/PLA/032	
Batch Number:	P700500	
Date of Manufacture:	20-DEC-16	
Expiry Date:	DEC-2018	

This is to certify that Crystel Platinum RTU solution above was tested on 20-DEC-16 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear Solution free of matter	Conforms
SPECIFIC GRAVITY	0.995 - 1.005	1.000
рH	8.50 – 10.00	9.3
STERILITY	No growth	Conforms

Completed By:	Reviewed By:
Quality Department: <u>Canokus</u>	Sapa C
Date: 10 Feb 2017.	Date: 10 FEB 17

DRF1500 CRY/PLA/033/Issue3/June 2012





Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY110 Platinum

Lot No:P700500 ICN: 0117004D

Subcontractor's Report Number: 2017009086

Lab Book Reference: ST03-56 Date Received: 12 JAN 2017 Date Completed: 08 FEB 2017

No. of samples: 10

Sterility Test - Membrane Filtration

Sterility Test - Membrane Filtration	
Medium	Result
Tryptone Soy Broth	Subculture
	Subculture
Fluid Thioglycollate Medium	

Subcultured as a precaution due to potential product precipitation

Medium	Result
TSA - 35 Degrees	No growth
TSA - 25 Degrees	No growth
13A = 23 Degrees	

Conclusions:

The samples submitted comply with the test for sterility in accordance with the client's methodology	
Results Reported by:	THE RESIDENCE OF THE PARTY OF T
Laboratory Technician: Charlotte Ray	Date: ©8 Feg 2017
Reviewed by:	
□ Quality Assurance Manager □ UK General Manager □ Technical Manager □ Laboratory Manager	Date: 08 (602017
QA Approved by:	
□ UK General Manager	Date: 09 F66 2017
☐ Technical Manager	Page 1 of
CONFIDENTIAL	Page 10

ST03-56 REVISION NO: 1

ICN: 0117004D DATE: 08 FEB 2017 TSCTS044-1 Rev 2

UK Deacon Park Moorgate Road Knowsky Liverpool L33 7RX Tel: +44 (0)151 547 7444 Pax: +44 (0)151 547 7400

124 Bernard F. Saint Jean Drive East Falmouth MA 02536-4445 USA "Tel: 001 508 540 3444 Fax: 001 508 540 8680

PYROQUANT DIAGNOSTIK GmbH Opelstrasse 14 D-64546 Mörfelden-Walldorf Germany Tet 0049 6105 96 10 0 Fac: 0049 6105 96 10 15



Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	C RAY	08 FEB 2017

CONFIDENTIAL

Page 2 of 2

ST03-56 REVISION NO: 1 ICN: 0117004D DATE: 08 FEB 2017

TSCTS044-1 Rev 2



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 25-May-2016

UK33S11630188-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1076147	
Customer Reference Number:	702371	
Product Description:	1000ml T/S DV4242 25-45kGy	
Validation Reference:	4242	
Quantity Received:	24	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	Bottle009 1000ml T/S Platinum, 1 Plt	
I	rradiation Data	
Date and Time of Irradiation:	25-May-2016 08:48	
Reference Dose Range kGy:	34.1 - 34.7	
Calculated Minimum Dose kGy:	32.3	
Calculated Maximum Dose kGy:	37.8	

Irradiation Release Authorised By Synergy Health plc



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 08-Sep-2015

UK33S11452158-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1087391	
Customer Reference Number:	700218	
Product Description:	Trigger spray head crystal-G DV4286 25-45kGy	
Validation Reference:	4286	
Quantity Received:	10	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	1plt	
1	rradiation Data	
Date and Time of Irradiation:	08-Sep-2015 04:53	
Reference Dose Range kGy:	31.1 - 32.1	
Calculated Minimum Dose kGy:	28.1	
Calculated Maximum Dose kGy:	34.8	

Irradiation Release Authorised By Synergy Health plc



http://www.synergyhealthplc.com

Certificate of Irradiation

Date issued: 19-Sep-2016 UK33S11702145-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1087393	
Customer Reference Number:	703090	
Product Description:	BAG 001-G DV4285 25-45kGy	
Validation Reference:	4285	
Quantity Received:	4	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	BAG001-G, 1PLT	
I	rradiation Data	
Date and Time of Irradiation:	19-Sep-2016 11:35	
Reference Dose Range kGy:	34.2 - 34.2	
Calculated Minimum Dose kGy:	26.4	
Calculated Maximum Dose kGy:	36.1	

Irradiation Release Authorised By Synergy Health plc