

Tristel

CERTIFICATE OF ANALYSIS		
PRODUCT NAME:	PERASTEL 3 DAY MANUAL WORKING SOLUTION	
REFERENCE:	MED/PER/037	
THIS IS TO CERTIFY THAT	Perastel 3 Day Manual Working Solution	
BEARING THE BATCH NUMBERS	P700900	
WAS TESTED ON	25-OCT-16	
AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS		
PROPERTIES	Specification	Results
APPEARANCE	Slightly hazy, pink liquid free of suspended matter	Conforms
SPECIFIC GRAVITY	1.015 +/- 0.010	1.013
pH	7.0 – 8.0	7.6
Completed By:	Reviewed By:	
Quality Department: <i>E. Andrews</i>	<i>GLD</i>	
Date: <i>11 Jan 2017</i>	Date: <i>11 JAN-17</i>	



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Perastel 3 Day Manual Base Solution

REFERENCE: MED/PER/034

DATE OF MANUFACTURE: 24-OCT-16

EXPIRY DATE: OCT-2018

THIS IS TO CERTIFY THAT Perastel 3 Day Manual Base Solution

BEARING THE BATCH NUMBER P700900

WAS TESTED ON 24-OCT-16

AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS

PROPERTIES	Specification	Results
APPEARANCE	Clear to pale yellow liquid free from haziness or suspended matter	Conforms
SPECIFIC GRAVITY	1.015 +/- 0.010	1.022
pH	7.0 - 8.0	7.7

Completed By:

Quality Department: *E Andrews*


Date: *11 Jan 2017*

Reviewed By:

[Signature]

Date: *11-JAN-17*



CERTIFICATE OF ANALYSIS		
PRODUCT NAME:	Perastel 3 Day Manual Activator Solution – UK1 & EX1	
REFERENCE:	MED/PER/035	
DATE OF MANUFACTURE:	10-OCT-16	
EXPIRY DATE:	OCT-2018	
THIS IS TO CERTIFY THAT	Perastel 3 Day Manual Activator Solution	
BEARING THE BATCH NUMBER	P700900	
WAS TESTED ON	11-OCT-16	
AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS		
PROPERTIES	Specification	Results
APPEARANCE	Pale red/pink viscous liquid with evenly distributed particulate matter	Conforms
SPECIFIC GRAVITY	1.020 +/- 0.010	1.015
pH	7.0 – 8.0	7.3
Completed By:	Reviewed By:	
Quality Department: <u>Canberra</u> 	
Date: <u>11 Jan 2017</u>	Date: <u>11 JAN - 17</u>	

