

Tristel

CERTIFICATE OF ANALYSIS

PRODUCT NAME: PERASTEL 3 DAY MANUAL WORKING SOLUTION

REFERENCE: MED/PER/037

THIS IS TO CERTIFY THAT Perastel 3 Day Manual Working Solution

BEARING THE BATCH NUMBERS P701005

WAS TESTED ON 10-JAN-17

AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS

PROPERTIES	Specification	Results
APPEARANCE	Slightly hazy, pink liquid free of suspended matter	Conforms
SPECIFIC GRAVITY	1.015 +/- 0.010	1.018
pH	7.0 - 8.0	7.6

Completed By:

Reviewed By:

Quality Department: E Andrews

.....

Date: 16 Jan 17

Date: 16 - JAN - 2017



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Perastel 3 Day Manual Base Solution

REFERENCE: MED/PER/034

DATE OF MANUFACTURE: 05-JAN-17

EXPIRY DATE: JAN-2019

THIS IS TO CERTIFY THAT Perastel 3 Day Manual Base Solution

BEARING THE BATCH NUMBER P701005

WAS TESTED ON 06-JAN-17

AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS

PROPERTIES	Specification	Results
APPEARANCE	Clear to pale yellow liquid free from haziness or suspended matter	Conforms
SPECIFIC GRAVITY	1.015 +/- 0.010	1.018
pH	7.0 - 8.0	7.8

Completed By:

Reviewed By:

Quality Department: *E Andrews*

Date: *16 Jan 17*

[Signature]
Date: *16-JAN-2017*



CERTIFICATE OF ANALYSIS		
PRODUCT NAME:	Perastel 3 Day Manual Activator Solution – UK1 & EX1	
REFERENCE:	MED/PER/035	
DATE OF MANUFACTURE:	03-JAN-17	
EXPIRY DATE:	JAN-2019	
THIS IS TO CERTIFY THAT	Perastel 3 Day Manual Activator Solution	
BEARING THE BATCH NUMBER	P701005	
WAS TESTED ON	04-JAN-17	
AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS		
PROPERTIES	Specification	Results
APPEARANCE	Pale red/pink viscous liquid with evenly distributed particulate matter	Conforms
SPECIFIC GRAVITY	1.020 +/- 0.010	1.017
pH	7.0 – 8.0	7.5
Completed By:	Reviewed By:	
Quality Department: <u>E. Andrews</u>		
Date: <u>16 Jan 17</u>	Date: <u>16-JAN-2017</u>	

