

Tristel

CERTIFICATE OF ANALYSIS

PRODUCT NAME: PERASTEL 3 DAY MANUAL WORKING SOLUTION

REFERENCE: MED/PER/037

THIS IS TO CERTIFY THAT Perastel 3 Day Manual Working Solution

BEARING THE BATCH NUMBERS P701100

WAS TESTED ON 11-JAN-17

AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS

PROPERTIES	Specification	Results
APPEARANCE	Slightly hazy, pink liquid free of suspended matter	Conforms
SPECIFIC GRAVITY	1.015 +/- 0.010	1.016
pH	7.0 - 8.0	7.6

Completed By:

Reviewed By:

Quality Department: E Andrews

Date: 16 Jan 17

.....

Date: 16-JAN-2017



Tristel

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Perastel 3 Day Manual Base Solution

REFERENCE: MED/PER/034

DATE OF MANUFACTURE: 09-JAN-17

EXPIRY DATE: JAN-2019

THIS IS TO CERTIFY THAT Perastel 3 Day Manual Base Solution

BEARING THE BATCH NUMBER P701100

WAS TESTED ON 09-JAN-17

AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS

PROPERTIES	Specification	Results
APPEARANCE	Clear to pale yellow liquid free from haziness or suspended matter	Conforms
SPECIFIC GRAVITY	1.015 +/- 0.010	1.020
pH	7.0 - 8.0	7.8

Completed By:

Reviewed By:

Quality Department: *E Andrews*

.....

Date: *16 Jan 17*

Date: *16 JAN 2017*



CERTIFICATE OF ANALYSIS

PRODUCT NAME: Perastel 3 Day Manual Activator Solution – UK1 & EX1

REFERENCE: MED/PER/035

DATE OF MANUFACTURE: 04-JAN-17

EXPIRY DATE: JAN-2019

THIS IS TO CERTIFY THAT Perastel 3 Day Manual Activator Solution

BEARING THE BATCH NUMBER P701100

WAS TESTED ON 05-JAN-17

AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS

PROPERTIES	Specification	Results
APPEARANCE	Pale red/pink viscous liquid with evenly distributed particulate matter	Conforms
SPECIFIC GRAVITY	1.020 +/- 0.010	1.016
pH	7.0 – 8.0	7.4

Completed By:

Reviewed By:

Quality Department: E Andrews

Date: 16 Jan 17

Date: 16-Jan-2017

