

## **CERTIFICATE OF ANALYSIS** PRODUCT NAME: PERASTEL 3 DAY MANUAL WORKING SOLUTION **REFERENCE:** MED/PER/037 THIS IS TO CERTIFY THAT Perastel 3 Day Manual Working Solution **BEARING THE BATCH NUMBERS** P701106 11-JAN-17 WAS TESTED ON AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS **PROPERTIES** Specification Results Slightly hazy, pink liquid free of **APPEARANCE** Conforms suspended matter SPECIFIC GRAVITY 1.015 +/- 0.010 1.017 pН 7.0 - 8.07.6 Completed By: Reviewed By: Quality Department: Con Oreus S ..... Date: 16 Jan 17

Document No: QCA-MED/PER/037-4

Page 1 of 1

Date: 16- JAN - 2017





CERTIFICATE OF ANALYSIS		
PRODUCT NAME:	Perastel 3 Day Manual Base Solution	
REFERENCE:	MED/PER/034	
DATE OF MANUFACTURE:	10-JAN-17	
EXPIRY DATE:	JAN-2019	
THIS IS TO CERTIFY THAT	Perastel 3 Day Manual Base Solution	
BEARING THE BATCH NUMBER P701106		
WAS TESTED ON	10-JAN-17	
AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS		
PROPERTIES	Specification	Results
APPEARANCE	Clear to pale yellow liquid free from haziness or suspended matter	Conforms
SPECIFIC GRAVITY	1.015 +/- 0.010	1.021
рН	7.0 – 8.0	7.7
Completed By:	Reviewed By:	
Quality Department: Canaleus		
Date: 16-Jan-17 Date: 16-Jan-2017		

Document No: QCA-MED/PER/034-4

Page 1 of

**DRF 1467** 



## **CERTIFICATE OF ANALYSIS** Perastel 3 Day Manual Activator Solution - UK1 & EX1 **PRODUCT NAME:** MED/PER/035 REFERENCE: DATE OF MANUFACTURE: 04-JAN-17 **EXPIRY DATE:** JAN-2019 Perastel 3 Day Manual Activator Solution THIS IS TO CERTIFY THAT BEARING THE BATCH NUMBER P701106 05-JAN-17 WAS TESTED ON AND FOUND TO COMPLY WITH ALL REQUIREMENTS AS SET OUT IN THE SPECIFICATIONS AND METHOD OF ANALYSIS Specification Results **PROPERTIES** Conforms Pale red/pink viscous liquid **APPEARANCE** with evenly distributed particulate matter 1.016 1.020 +/- 0.010 SPECIFIC GRAVITY 7.0 - 8.07.4 Ηq Reviewed By: Completed By: Quality Department: CONDIGUES ..... Date: 16 Jan 17 Date: 16-JAN-2017

Document No: QCA-MED/PER/035-5

