







Certificate of Analysis-TITANIUM WIPES-CRY/TTM/049

PRODUCT batch number: P701900

Product Name:		TITANIUM WIPES Working Solution			
Reference:		CRY/TTM/047			
This is to certify that TITANIU	M WIPES Wo	rking Solution b	earing the Batch No's:		
P701900 were tested on 08- set out in the specifications		• •	with all requirements as		
PROPERTIES	SPECIFICATION		RESULTS		
APPEARANCE	Clear pale yellow liquid free of matter				Conforms
ODOUR	Chlorine like		Conforms		
ClO2 Concentration (ppm)	Conforms to QCF CRY/TTM/050 Conforms		Conforms		

Product Name:		TITANIUM WIPES Base Solution		
Reference:		CRY/TTM/045		
Date of Manufacture:		08-DEC-16		
Date of Expiry:		DEC-2018		
This is to certify that TIT DEC-16 and found to co and method of analysis	mply with all requ			
PROPERTIES	SPECIFICATION		RESULTS	
APPEARANCE	Clear liquid free of matter		Conforms	
SPECIFIC GRAVITY	1.000-1.010		1,006	
рН	2.0 - 3.0		2.2	
STERILITY	No growth		Conforms	

DRF2550 CRY/TTM/049/Issue1











Product Name:	TITANIUM WIPES Activator Solution
Reference:	CRY/TTM/046
Date of Manufacture:	01-DEC-16
Date of Expiry:	DEC-2018
This is to certify that TITANIUM \	NIPES Activator Solution 633602 was tested on 01-
	th all requirements as set out in the specifications

and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear and free of matter	Conforms
рН	11.0-11.5	11.1
SODIUM CHLORITE	0.2% +/- 0.005%	0.197%
STERILITY	No growth	Conforms

Completed By:	Reviewed By:
Quality Department: <u>CODDRUS</u>	TEMPED.
Date: 01-MARCH 2017	Date:O. MAR 17

DRF2550 CRY/TTM/049/Issue1





Tristel Solutions Ltd Unit 1B, Lynx Business Park Fordham Road Cambridgeshire CB8 7NY

Certificate of Analysis

Sample Details Product Name: Titanium burstable wipes

Batch Number:

P701900

Batch Code P701900

Job Number:

15991

Sample ID:

56899

Registered date:

14-Feb-17

Completion date:

01-Mar-17

Data

Test Method	Test	Specification	Result	Units	Pass/Fail
Direct Inoculation Sterility Test (cFTM)	DI Sterility (cFTM)	Complies / Does Not Cornply	Complies	N/A	Pass
Direct Inoculation Sterility Test (TSB)	DI Sterility (TSB)	Complies / Does Not Comply	Complies	N/A	Pass
Membrane Filtration Sterility Test (cFTM)	MF Sterility (cFTM)	Complies / Does Not Cornply	Complies	N/A	Pass
Membrane Filtration Sterility Test (TSB)	MF Sterility (TSB)	Complies / Does Not Cornply	Complies	N/A	Pass

The product meets the following specification:

Sterility Test current Ph. Eur. by Membrane Filtration and Direct Inoculation. Validation report HVAL_R405.v1

LabHQ Product Specification Version:

Approval

Approved By

Karen Nash

Approved Date:

01-Mar-2017 10:53:24

User ID

knash

Title

General Manager

This approval is performed by electronic signature in LabHQ LIMS and is the legal equivalent of the approver's hand-written signature.



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 28-Jan-2017 UK31S11796341-1-1

This is to certify that Elgin Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Or	der information
Account Number: Synergy Health Sales Part Reference: Customer Reference Number: Product Description: Validation Reference: Quantity Received: Customer Minimum Specification kGy: Customer Maximum Specification kGy: Customer Unit Lot/Batch Number:	100708 1103783 704050 BURST/TITANIUM WIPES 10-20KGY 1.6922 40 10.0 20.0 CRY 144, LOT NO. P701900, 1 PALLET
li li	rradiation Data
Date and Time of Irradiation: Reference Dose Range kGy: Calculated Minimum Dose kGy: Calculated Maximum Dose kGy:	28-Jan-2017 00:28 14.0 - 14.4 11.4 14.5

Irradiation Release Authorised By Synergy Health plc