



Certificate of Analysis-CRY 112

PRODUCT batch number: P715702
 BASE batch number: P715701
 ACTIVATOR batch number: P715700

Product Name:	Kryptonite Working Solution	
Product batch number:	P715702	
Reference:	CRY/SPO/021	
This is to certify that Kryptonite Working Solution bearing the batch no's: P715701 and P715700 on 13-Mar-2017 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear pale yellow liquid free of matter	Conforms
SPECIFIC GRAVITY	0.995 – 1.005	1.000
pH	2.5 – 5.5	3.1
ClO ₂ Concentration – after 10 minutes (ppm)	230 – 280	270

Product Name:	Kryptonite Base Solution	
Reference:	CRY/SPO/020	
Date of Manufacture:	13-Mar-2017	
Date of Expiry:	Mar-2019	
This is to certify that Kryptonite Base Solution was tested on 13-Mar-2017 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear blue liquid free of matter	Conforms
SPECIFIC GRAVITY	1.020 – 1.060	1.022
pH	1.5 – 3.5	2.0
STERILITY	No growth	Conforms





Product Name:	Kryptonite Activator Solution	
Reference:	CRY/SPO/019	
Date of Manufacture:	13-Mar-2017	
Date of Expiry:	Mar-2019	
This is to certify that Kryptonite Activator Solution was tested on 13-Mar-2017 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear and free of matter	Conforms
SPECIFIC GRAVITY	1.015 – 1.035	1.015
pH	11.5 – 13.0	11.8
SODIUM CHLORITE	2.10% +/- 0.1%	2.13%
STERILITY	No growth	Conforms

Completed By:

Quality Department:

Date: 2017-06-29

Reviewed By:

..... E Andrews

Date: 29 Jun 2017



Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY112 Kryptonite Base	Lab Book Reference: ST03-85
Lot No: P715701	Date Received: 09 Jun 2017
ICN: 0617006D	Date Completed: 29 Jun 2017
Subcontractor's Report Number: 2017104505	No. of samples: 10

Sterility Test – Membrane Filtration

Medium	Result
Tryptone Soy Broth	No growth
Fluid Thioglycollate Medium	No growth

Conclusions:

The samples submitted comply with the test for sterility in accordance with the client's methodology.

Results Reported by:

S.Hughes
.....
Laboratory Technician/Designee: Sophie Hughes

Date: 29 Jun 2017

Reviewed by:

M. [Signature]
.....
 Quality Assurance Manager
 UK General Manager
 Technical Manager
 Laboratory Manager

Date: 29 Jun 2017

QA Approved by:

M. [Signature]
.....
 Quality Assurance
 UK General Manager
 Technical Manager

Date: 29 Jun 2017

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	S. Hughes	29 Jun 2017

CONFIDENTIAL

ST03-85ICN: 0617006D

Page 1 of 1
TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 29 Jun 2017

UK
Deacon Park
Moorgate Road
Knowsley
Liverpool L33 7RX
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA
124 Bernard I, Saint Jean Drive
East Falmouth
MA 02536-4445
USA
Tel: 001 508 540 3444
Fax: 001 508 540 8680

Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Morfelden-Walldorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15

Web: www.acciuk.co.uk

Registered in England and Wales. Company Registration Number: BR002906

Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY112 Kryptonite Activator	Lab Book Reference: ST03-84
Lot No: P715700	Date Received: 09 Jun 2017
ICN: 0617006D	Date Completed: 29 Jun 2017
Subcontractor's Report Number: 2017104504	No. of samples: 10

Sterility Test - Membrane Filtration

Medium	Result
Tryptone Soy Broth	No growth
Fluid Thioglycollate Medium	No growth

Conclusions:

The samples submitted comply with the test for sterility in accordance with the client's methodology.

Results Reported by:


.....
Laboratory Technician/Designee: Sophie Hughes


Date: 29 JUN 2017

Reviewed by:


.....
 Quality Assurance Manager
 UK General Manager
 Technical Manager
 Laboratory Manager

Date: 29 Jun 2017

QA Approved by:


.....
 Quality Assurance
 UK General Manager
 Technical Manager

Date: 29 Jun 2017

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	S. Hughes	29 Jun 2017

CONFIDENTIAL

ST03-84ICN: 0617006D

Page 1 of 1
TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 29 Jun 2017

UK
Dragon Park
Moorgate Road
Knuwsley
Liverpool L33 7RN
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA
124 Bernard E. Saint Jean Drive
East Falmouth
MA 02536-4445
USA
Tel: 001 508 540 3444
Fax: 001 508 540 8680

Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Mörfelden-Walldorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15

Web: www.acciuk.co.uk

Registered in England and Wales. Company Registration Number: BR002906