







Certificate of Analysis

Product Name:	Crystel Platinum RTU Solution	
Reference:	CRY/PLA/032	
Batch Number:	P720500	
Date of Manufacture:	01-FEB-17	
Expiry Date:	FEB-2019	

This is to certify that Crystel Platinum RTU solution above was tested on 03-FEB-17 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear Solution free of matter	Conforms
PECIFIC GRAVITY	0.995 - 1.005	0.999
Н	8.50 - 10.00	9.2
STERILITY	No growth	Conforms

Completed By:

Quality Department: (= Chookeus)

Date: 2017-08-23

Reviewed By:

Date: 2017-08-23

DRF1500 CRY/PLA/033/Issue3/June 2012





Specialists in Endotoxin and Glucan Detection

Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY110 Platinum

Lot No: P720500 ICN: 0717018D

Subcontractor's Report Number: 2017138636

Lab Book Reference: ST03-86 Date Received: 25 Jul 2017 Date Completed: 16 Aug 2017

No. of samples: 10

Sterility Test - Membrane Filtration

occimely rest Pichibiane includes	
Medium	Result
Tryptone Soy Broth	No Growth
Fluid Thioglycollate Medium	No Growth

Conclusions:

The samples submitted comply with the test for sterility	in accordance with the client's methodology
--	---

The samples submitted comply with the test for sterility in accordance	e with the client's methodology.
Results Reported by:	
Laboratory Technician/Designee: Sophie Hughes	Date: 16 AUG 2017
Reviewed by:	
☐ Quality Assurance Manager ☐ UK General Manager ☐ Technical Manager ☐ Caboratory Manager	Date: 18 Aug 2017
QA Approved by:	
(Mellinger)	Date: 18 Aug 2017
☑ Quality Assurance ☑ UK General Manager ☑ Technical Manager	

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	S. Hughes	16 Aug 2017

CONFIDENTIAL ST03-86

ICN: 0717018D

Page 1 of 1 TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 16 Aug 2017

UK Deacon Park Moorgate Road Knowley Liverpool L33 7RX Tel: +44 (0)151 547 7444 Fax: +44 (0)151 547 7400

124 Bernard E, Saint Jean Drive East Falmouth MA 02536-1445 USA Tel: 001 508 540 3444 Fax: 001 508 540 8680

Germany PYROQUANT DIAGNOSTIK GmbH Opelstrasse 14 D-64546 Mörfelden-Walldorf Germany Tel: 0049 6105 96 10 0 Fax: 0049 6105 96 10 15



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 23-Jul-2016 UK33S11669509-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd Unit 4C Lynx Bus Park, Fordham Rd Newmarket Suffolk CB8 7NY UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1076147	
Customer Reference Number:	702752	
Product Description:	1000ml T/S DV4242 25-45kGy	
Validation Reference:	4242	
Quantity Received:	49	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	BOTTLE 009-PLATINUM, 2 PLTS	
I	rradiation Data	
Date and Time of Irradiation:	23-Jul-2016 07:35	
Reference Dose Range kGy:	31.6 - 32.2	
Calculated Minimum Dose kGy:	30.0	
Calculated Maximum Dose kGy:	35.0	

Irradiation Release Authorised By Synergy Health plc



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 08-Sep-2015 UK33S11455485-1-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd Unit 4C Lynx Bus Park, Fordham Rd Newmarket Suffolk CB8 7NY UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1087391	
Customer Reference Number:	700363	
Product Description:	Trigger spray head crystal-G DV4286 25-45kGy	
Validation Reference:	4286	
Quantity Received:	15	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	1pit	
I	rradiation Data	
Date and Time of Irradiation:	08-Sep-2015 04:54	
Reference Dose Range kGy:	31.3 - 32.0	
Calculated Minimum Dose kGy:	28.3	
Calculated Maximum Dose kGy:	34.7	

Irradiation Release Authorised By Synergy Health plc



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 29-Aug-2016 UK33S11678232-3-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1087393	
Customer Reference Number:	702828	
Product Description:	BAG 001-G DV4285 25-45kGy	
Validation Reference:	4285	
Quantity Received:	3	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	BAG 001-G, 1 plt	
	rradiation Data	
Date and Time of Irradiation:	29-Aug-2016 05:43	
Reference Dose Range kGy:	37.6 - 37.6	
Calculated Minimum Dose kGy:	29.0	
Calculated Maximum Dose kGy:	39.6	

Irradiation Release Authorised By Synergy Health plc