







Certificate of Analysis-TITANIUM -CRY151

PRODUCT batch number: P729200

Product Name:	TITANIUM Work	TITANIUM Working Solution	
Reference:	CRY/TTM/024		
	UM Working Solution bearing the Batch No's: P729200 and found to comply with all requirements as set out in the bod of analysis.		
PROPERTIES	SPECIFICATION	RESULTS	
APPEARANCE	Clear pale yellow liquid free of matter	Conforms	
ODOUR	Chlorine like	Conforms	
ClO ₂ Concentration (ppm)	160-200	192	

Product Name:		TITANIUM Base	Solution
Reference:		MED/SOL/449	
Date of Manufacture:		2017-10-10	
Date of Expiry:		2019-10	
This is to certify that TIT. to comply with all requiremanalysis.			-
PROPERTIES	SPECIF	ICATION	RESULTS
APPEARANCE	Clear liquid free of matter		Conforms
SPECIFIC GRAVITY	1.000-1.200		1.022
pH	1.5 – 3.5		1.9
STERILITY	No growth		Conforms

CRY/TTM/007/Issue5

DRF2616











Product Name:		TITANIUM Activator Solution	
Reference:	N	ED/SOL/450	
Date of Manufacture:	2	017-10-10	
Date of Expiry:		019-10	
This is to certify that TITA found to comply with all of analysis.	requirements as se	t out in the sp	ecifications and method
PROPERTIES	SPECIFICA	TION	RESULTS
APPEARANCE	Clear and free	of matter	Conforms
pH	10.0-11	.0	11.0
SODIUM CHLORITE	0.1% +/- 0	.005%	1.002
STERILITY	No grov	vth	Conforms

Completed By:	Reviewed By:
Quality Department: CONDIGUES	- Rielin Xa
Date: 2017 - 11 - 214	Date: 2017-11-24

CRY/TTM/007/Issue5

DRF2616





Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: Cry151 Titanium Base

Lot No: P729200 ICN: 1017021D

Subcontractor's Report Number: 2017199633

Lab Book Reference: ST03-112 Date Received: 24 Oct 2017 Date Completed: 17 Nov 2017

No. of samples: 10

Sterility Test - Membrane Filtration

Oterine, rest Tremerane intraction	
Medium	Result
Tryptone Soy Broth	No Growth
Fluid Thioglycollate Medium	No Growth

Conclusio The samples		omply with the test for sterility in	accordance with the clie	nt's methodology.
	echnician/Design	nee: C.Thornton	Date: 171	wort
Quality Assu UK General I Technical Ma	rance Manager Manager anager)	Date: 17 N	cv 2017
QA Appro Quality Assu UK General I Technical Ma	rance Manager		Date:	Nov2017
Revision I	History:			
DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	C.Thornton	17 Nov 2017

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	C.Thornton	17 Nov 2017

CONFIDENTIAL

ST03-112ICN: 1017021D

Page 1 of 1 TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 17 Nov 2017

UK Deacon Park Moorgate Road Knowsley Everpool L33 7RX Tel: +44 (0)151 547 7144 Pax +44 (0)151 547 7400

USA 124 Bernard F. Saint Jean Urive East Falmouth MA 02536-4445 Tel: 001 508 540 3444 Fax: 001 508 540 8680

Germany PYROQUANT DIAGNOSTIK GmbH Opelstrasse 14 D-64546 Morfelden-Walldorf

Tel 0049 6105 96 10 0 Fac: 0049 6105 96 10 15



Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: Cry151 Titanium Activator

Lot No: P729200 ICN: 1017021D

Subcontractor's Report Number: 2017199634

Lab Book Reference: ST03-114 Date Received: 24 Oct 2017 Date Completed: 17 Nov 2017

No. of samples: 10

Sterility Test - Membrane Filtration

Medium	Result
Tryptone Soy Broth	No Growth
Fluid Thioglycollate Medium	No Growth

Conclusio				
		omply with the test for sterility in	accordance with the cl	ient's methodology
ine sumple	. J Jubillitied C	ompry with the test for stermey in	accordance with the ci	iches methodology.
Results R	eported by	:		
Laboratory T	echnician/Desig	nee: C.Thornton	Date:	rknsort
Quality Assu	urance Manager Manager Janager		Date:	va 2017
QA Appro	urance Manager		Date: /7	No/2017
Revision I	History:			
DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	C.Thornton	17 Nov 2017

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	C.Thornton	17 Nov 2017

CONFIDENTIAL

ST03-114ICN: 1017021D

Page 1 of 1 TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 17 Nov 2017

UK Deacon Park Moorgate Road Knowsky Everpool 1.33 7RX Tel +44 (0)151 547 7444 Tax: +44 (0)151 547 7400

124 Bernard F. Saint Jean Dove East Falmouth MA 02536-4145 USA Tel 001 508 540 3444 Fax: 001 508 540 8680

Germany PYROQUANT DIAGNOSTIK GmbH Opelstrasse 14 D-64546 Morfelden-Walldort Tel: 0049 6105 96 10 0 Jax 0049 6105 96 10 15

Web; www.acciuk.co.uk



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 05-Oct-2017 UK33S11974005-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Order Information			
Account Number:	100708		
Synergy Health Sales Part Reference:	1077586		
Customer Reference Number:	705659		
Product Description:	JET TITANIUM DV4260 25-40kGy		
Validation Reference:	4260		
Quantity Received:	16		
Customer Minimum Specification kGy:	25.0		
Customer Maximum Specification kGy:	40.0		
Customer Unit Lot/Batch Number:	BN: 20170921, 1 plt		
ı	rradiation Data		
Date and Time of Irradiation:	05-Oct-2017 05:19		
Reference Dose Range kGy:	33.9 - 34.1		
Calculated Minimum Dose kGy:	31.9		
Calculated Maximum Dose kGy:	37.0		



http://www.synergyhealthplc.com

Certificate of Irradiation

Date issued: 05-Oct-2017 UK33S11974005-2-3

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1077586	
Customer Reference Number:	705659	
Product Description:	JET TITANIUM DV4260 25-40kGy	
Validation Reference:	4260	
Quantity Received:	16	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	40.0	
Customer Unit Lot/Batch Number:	BN: 20170921, 1 plt	
I	rradiation Data	
Date and Time of Irradiation:	05-Oct-2017 05:19	
Reference Dose Range kGy:	33.6 - 34.0	
Calculated Minimum Dose kGy:	31.6	
Calculated Maximum Dose kGy:	36.9	

UK33S11930917-1-1



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 03-Aug-2017

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1077585	
Customer Reference Number:	705262	
Product Description:	JET TRIGGER HEAD DV4258 25-40kGy	
Validation Reference:	4258	
Quantity Received:	32	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	40.0	
Customer Unit Lot/Batch Number:	Lot 25072017, 2 pits	
I	rradiation Data	
Date and Time of Irradiation:	03-Aug-2017 12:23	
Reference Dose Range kGy:	31.1 - 31.6	
Calculated Minimum Dose kGy:	28.0	
Calculated Maximum Dose kGy:	34.5	



http://www.synergyhealthplc.com

Certificate of Irradiation

Date Issued: 11-Mar-2017 UK33S11824859-2-1

This is to certify that AST Daventry Synergy Health PLC has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products EN ISO 9001:2008 Quality Management System EN ISO 13485:2012 Quality System - Medical Devices

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1087393	
Customer Reference Number:	704309	
Product Description:	BAG 001-G DV4285 25-45kGy	
Validation Reference:	4285	
Quantity Received:	10	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	21022017, 1 plt	
I	rradiation Data	
Date and Time of Irradiation:	11-Mar-2017 06:48	
Reference Dose Range kGy:	33.7 - 34.3	
Calculated Minimum Dose kGy:	26.0	
Calculated Maximum Dose kGy:	36.2	