







### **Certificate of Analysis**

Product Name:	Crystel Platinum CRY109
Reference:	CRY/PLA/019
Batch Number:	P803900
Date of Manufacture:	2018-01-15
Expiry Date:	2020-01-15

This is to certify that Crystel Platinum solution above was tested on 2018-01-16 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear solution free of matter	Conforms
SPECIFIC GRAVITY	0.900 - 1.100	0.988
pH (neat)	9.0 – 10.0	9.9
STERILITY	No growth	No Growth

Completed By:	Reviewed By:
Quality Department: (Carolicus	Kelinso
Date: 2018-03-09.	Date: 2018-03-09

DRF1399 CRY/PLA/007/Issue3/June 2012





Specialists in Endotoxin and Glucan Detection

#### **Sterility Testing Analytical Report**

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY109 Platinum

Lot No: P803900 ICN: 0218005D

Subcontractor's Report Number: 2018029033

Lab Book Reference: ST03-132 Date Received: 07 Feb 2018 Date Completed: 07 Mar 2018

No. of samples: 10

Sterility Test - Membrane Filtration

Sternity rest - Membrane Fittation	
Medium	Result
Tryptone Soy Broth	No growth
Fluid Thioglycollate Medium	No growth

#### Conclusions:

The samples submitted comply with the test for sterility in accordance with the client's methodology.

The samples submitted samply man are seen and	
Results Reported by:	
Laboratory Technician/Designee: Daniel Jennings	Date: 07 MAR 2018
Reviewed by:	Date: 07 MAC 2018
☐ Quality Assurance Manager ☑ UK General Manager ☐ Technical Manager ☐ Laboratory Manager	
QA Approved by:	
alobalo	Date: 07 MAR2018
☐ Quality Assurance ☐ UK General Manager ☐ Technical Manager	

### **Revision History:**

		74		1
DCCF	No: Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	D. Jennings	07 Mar 2018

CONFIDENTIAL ST03-132

ICN: 0218005D

Page 1 of 1 TSCTS044-1 Rev 3

**REVISION NO: 1** 

DATE: 07 Mar 2018

UK
Deacon Park
Moorgate Road
Kniwsley
Läverpool L33 7RX
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA 124 Bernard F, Saint Jean Drive East Falmouth AtA 02536-4445 USA Tel: 001 508 540 3444 Fax: 001 508 540 8680 Germany PYROQUANT DIAGNOSTIK GmbH Opelstrasse 14 D-64546 Morfelden-Walldorf Germany Tel: 0049 6105 96 10 0 East 0049 6105 96 10 15



http://www.steris-ast.com

### **Certificate of Irradiation**

Date Issued: 11-Jan-2018

UK33S12034059-1-2

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1121497	
<b>Customer Reference Number:</b>	706219	
Product Description:	BOTTLE 100ML DV5155 25-45kGy	
Validation Reference:	5155	
Quantity Received:	16	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	14122017, 1 plt	
1	rradiation Data	
Date and Time of irradiation:	10-Jan-2018 21:39	
Reference Dose Range kGy:	32.7 - 33.1	
Calculated Minimum Dose kGy:	30.7	
Calculated Maximum Dose kGy:	35.9	

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company



http://www.steris-ast.com

# **Certificate of Irradiation**

Date Issued: 08-Jan-2018 UK33S12037706-1-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1121930	
Customer Reference Number:	706256	
Product Description:	100ML BOTTLE CAP DV5165 25-45kGy	
Validation Reference:	5165	
Quantity Received:	2	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	CAP001-G, BN: 20122017, EXP: 20191230, 1 pit	
lı	rradiation Data	
Date and Time of Irradiation:	08-Jan-2018 11:20	
Reference Dose Range kGy:	34.4 - 34.4	
Calculated Minimum Dose kGy:	31.6	
Calculated Maximum Dose kGy:	38.3	

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company



http://www.steris-ast.com

# **Certificate of Irradiation**

Date Issued: 15-Jan-2018

UK33S12041486-1-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1121936	
Customer Reference Number:	706279	
Product Description:	BAG 004-G DV5175 25-45kGy	
Validation Reference:	5175	
Quantity Received:	4	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	06112017, 1 PLT	
lı l	rradiation Data	
Date and Time of Irradiation:	15-Jan-2018 21:52	
Reference Dose Range kGy:	33.7 - 33.7	
Calculated Minimum Dose kGy:	25.9	
Calculated Maximum Dose kGy:	36.1	

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company