







# **Certificate of Analysis**

Product Name:	Crystel Platinum RTU Solution CRY1	
Reference:	CRY/PLA/032	
Batch Number:	P804001	
Date of Manufacture:	2018-01-17	
Expiry Date:	2020-01-17	

This is to certify that Crystel Platinum RTU solution above was tested on 2018-01-18 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear Solution free of matter	Conforms
SPECIFIC GRAVITY	0.995 - 1.005	1.000
рН	8.50 - 10.00	9.22
STERILITY	No growth	No growth

Completed By:	Reviewed By:		
Quality Department: Carolows	Sielinka		
Date: 2018-03-09	Date: 2018-03-09		

DRF1500 CRY/PLA/033/issue3/June 2012





#### **Sterility Testing Analytical Report**

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY110 Platinum

Lot No: P804001 ICN: 0218009D

**Subcontractor's Report Number**: 2018029071

Lab Book Reference: ST03-134 Date Received: 13 Feb 2018 Date Completed: 07 Mar 2018

No. of samples: 10

Manchuana Eiltration

Sternity lest - Membrane Filtration		
Medium	Result	
Tryptone Soy Broth	No growth	
Fluid Thioglycollate Medium	No growth	

#### **Conclusions:**

The samples	s submitted co	mply with the test for sterility in	accordance with the clie	ent's methodology.
Results Re	eported by:			
		nee: Daniel Jennings	Date: O.7	MAR ZOIS
Quality Assu	rance Manager Manager		Date: 67.1	une 2018
QA Appro  QA Appro  Quality Assu  UK General I  Technical Ma	ved by:  rance Manager	***************************************	Date:	MAR 8018
Revision l	History:			
DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	D. Jennings	07 Mar 2018

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	D. Jennings	07 Mar 2018

CONFIDENTIAL ST03-134

ICN: 0218009D

Page 1 of 1 TSCTS044-1 Rev 3

**REVISION NO: 1** 

**DATE: 07 Mar 2018** 

UK Deacon Park Moorgate Road Knowstey Liverpool L33 7RX lek +44 (0)151 547 7414

Fax +44 (0)151 547 7400

124 Bernard E. Saint Jean Drive East Falmouth MA 02536-4445 USA Tel: 001 508 540 3444 Fax: 001 508 540 8680

PYROQUANT DIAGNOSTIK GmbH Opelstrasse 11 D-64546 Marfeklen-Walldorf Germany Tel: 0049 6105 96 10 0 Fac: 0049 5105 96 10 15



http://www.steris-ast.com

# **Certificate of Irradiation**

Date Issued: 18-Jan-2018 UK33S12041486-4-2

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information			
Account Number: Synergy Health Sales Part Reference: Customer Reference Number: Product Description: Validation Reference: Quantity Received: Customer Minimum Specification kGy: Customer Maximum Specification kGy:	100708 1121939 706279 1000ML T/S BOTTLE DV5171 25-45kGy 5171 16 25.0 45.0		
Customer Unit Lot/Batch Number:	02112017, 1 PLT		
li li	rradiation Data		
Date and Time of Irradiation: Reference Dose Range kGy:	18-Jan-2018 17:24 30.2 - 30.4		
Calculated Minimum Dose kGy: Calculated Maximum Dose kGy:	29.0 33.1		

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company



http://www.steris-ast.com

## **Certificate of Irradiation**

Date Issued: 11-Jan-2018

UK33S12037910-2-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1121921	
Customer Reference Number:	706246	
Product Description:	TRIGGER SPRAY HEAD CRYSTAL-G DV5170	
	25-45kGy	
Validation Reference:	5170	
Quantity Received:	8	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	B/N: 21122017, EXP: 20191230, 1 plt	
I	rradiation Data	
Date and Time of Irradiation:	11-Jan-2018 07:30	
Reference Dose Range kGy:	31.3 - <mark>3</mark> 1.4	
Calculated Minimum Dose kGy:	28.5	
Calculated Maximum Dose kGy:	34.7	

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company



http://www.steris-ast.com

#### **Certificate of Irradiation**

Date Issued: 07-Feb-2018

UK33S12051056-1-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1121938	
Customer Reference Number:	706435	
Product Description:	BAG 001-G DV5194 25-48kGy	
Validation Reference:	5194	
Quantity Received:	5	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	48.0	
Customer Unit Lot/Batch Number:	20171017, 1 PLT	
I	rradiation Data	
Date and Time of Irradiation:	06-Feb-2018 18:57	
Reference Dose Range kGy:	38.4 - 38.4	
Calculated Minimum Dose kGy:	27.5	
Calculated Maximum Dose kGy:	42.8	

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company