







Certificate of Analysis-TITANIUM WIPES-CRY/TTM/049

PRODUCT batch number: P8o58oo

Product Name:		TITANIUM WIPES Working Solution		
Reference:		CRY/TTM/047		
This is to certify that TITANIUM WIPES Working Solution bearing the Batch No's:			earing the Batch No's:	
P805800 were tested on 2018.01.11 and found to comply with all requirements			with all requirements as	
set out in the specifications	and method	of analysis.		
PROPERTIES	SPECIFICATION		RESULTS	
APPEARANCE	Clear pale yellow liquid free of matter		Conforms	
ODOUR	Chlorine like		Conforms	
ClO2 Concentration (ppm)	Conforms to QCF CRY/TTM/050		Conforms	

Product Name:		TITANIUM WIPES Base Solution		
Reference:		CRY/TTM/045		
Date of Manufacture:		2018-01-10		
Date of Expiry:		2020-01-10		
This is to certify that TITANIUM Wipes Base Solution was tested on 2018-01-11 and found to comply with all requirements as set out in the specifications and method of analysis.				
PROPERTIES	SPECIFICATION		RESULTS	
APPEARANCE	Clear liquid free of matter		Conforms	
SPECIFIC GRAVITY	1.000-1.010		1.005	
рН	2.0 – 3.0		2.2	
STERILITY	No growth		Conforms	

DRF2550 CRY/TTM/049/Issue1











Product Name:	TITANIUM W	TITANIUM WIPES Activator Solution		
Reference:	CRY/TTM/oz	CRY/TTM/046		
Date of Manufacture:	2017-10-03	2017-10-03		
Date of Expiry:	2019-10-03	2019-10-03		
•	ANIUM WIPES Activator Solution in all requirements as set out i			
PROPERTIES	SPECIFICATION	RESULTS		
APPEARANCE	Clear and free of matter	Conforms		
рН	11.0-11.5	11.1		
SODIUM CHLORITE	0.2% +/- 0.005%	0.198%		
STERILITY	No growth	Conforms		

Completed By: Quality Department:		Reviewed By:		
		Skielinso		
Date:	2018-05-22	Date: 2018-05-22		

DRF2550 CRY/TTM/049/issue1







P805800

Certificate of Analysis

Titanium burstable wipes **Sample Details Product Name:**

Job Number: 22130 P805800 **Batch Code** 76802 Sample ID:

> 04-May-18 Registered date:

Completion date: 21-May-18

Data

Batch Number:

Test Method	Test	Specification	Result	Units	Pass/Fail
Direct Inoculation Sterility Test (cFTM)	DI Sterility (cFTM)	Complies / Does Not Cornply	Complies	N/A	Pass
Direct Inoculation Sterility Test (TSB)	DI Sterility (TSB)	Complies / Does Not Cornply	Complies	N/A	Pass
Membrane Filtration Sterility Test (cFTM)	MF Sterility (cFTM)	Complies / Does Not Cornply	Complies	N/A	Pass
Membrane Filtration Sterility Test (TSB)	MF Sterility (TSB)	Complies / Does Not Cornply	Complies	N/A	Pass

The product meets the following specification:

Sterility Test current Ph. Eur. by Membrane Filtration and Direct Inoculation. Validation report HVAL_R405.v2

LabHQ Product Specification Version:

Approval

21-May-2018 11:31:23 **Approved Date:** Katie Pearson **Approved By**

kpearson **User ID**

Laboratory Supervisor Title

This approval is performed by electronic signature in LabHQ LIMS and is the legal equivalent of the approver's hand-written signature.



http://www.steris-ast.com

Certificate of Irradiation

Date Issued: 10-Apr-2018

UK31S12071669-1-1

This is to certify that Elgin Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information			
Account Number: Synergy Health Sales Part Reference: Customer Reference Number: Product Description: Validation Reference: Quantity Received: Customer Minimum Specification kGy: Customer Maximum Specification kGy: Customer Unit Lot/Batch Number:	100708 1103783 706711 BURST/TITANIUM WIPES 10-20KGY 1.6922 54 10.0 20.0 LOT P805800, 2 PALLETS		
lı	rradiation Data		
Date and Time of Irradiation: Reference Dose Range kGy: Calculated Minimum Dose kGy: Calculated Maximum Dose kGy:	10-Apr-2018 10:01 17.8 - 17.9 14.9 18.0		

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company