







Certificate of Analysis

Product Name:	Crystel Platinum RTU Solution CRY110
Reference:	CRY/PLA/032
Batch Number:	P807500
Date of Manufacture:	2018-03-08
Expiry Date:	2020-03-08

This is to certify that Crystel Platinum RTU solution above was tested on 2018-03-09 and found to comply with all requirements as set out in the specifications and method of analysis.

PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear Solution free of matter	Conforms
SPECIFIC GRAVITY	0.995 - 1.005	1.000
рН	8.50 – 10.00	9.3
STERILITY	No growth	Conforms

Complet	ed By:	Reviewed By:
Quality [Department: Cand Veus	Kielinka
Date:	2018-04-20.	Date: 2018-04-26

DRF1500 CRY/PLA/033/Issue3/June 2012





Specialists in Endotoxin and Glucan Detection

Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY110 Platinum

Lot No: P807500 ICN: 0318020F

Subcontractor's Report Number: 2018054189

Lab Book Reference: ST03-152

Date Received: 22 Mar 2018 Date Completed: 17 Apr 2018

No. of samples: 10

Sterility Test - Membrane Filtration

Training rest remarkable restation	
Medium	Result
Tryptone Soy Broth	No growth
Fluid Thioglycollate Medium	No growth

Conclusions:

Results Reported by:

The samples submitted comply with the test for sterility in accordance with the client's methodology.

Laboratory Technician/Designee: Daniel Jennings	Date: 13 APR 2018
Reviewed by:	- 100 A 110
Quality Assurance Manager UK General Manager Technical Manager Laboratory Manager	Date: MARC 1018
QA Approved by:	
Ulskouls	Date:
Quality Assurance UK General Manager Technical Manager	

Revision History:

DC	CF No:	Revision:	Description:	Originator:	Effective Date:
N/	Α	1	Revision 1, no changes	D. Jennings	17 Apr 2018

CONFIDENTIAL ST03-152

ICN: 0318020F

Page 1 of 1 TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 17 Apr 2018

UK Deacon Park Mongate Road Knowsky Liverpool LS3 7RN Tel. +44 (0)151 547 7444 Fax: +44 (0)151 547 7400 USA 124 Bernard E. Saint Jean Drive East Falmouth MA 02536-1445 USA Tel: 001 508 540 3444 Fax: 001 508 540 8680 Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Mörfelden-Walldorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15



http://www.steris-ast.com

Certificate of Irradiation

Date Issued: 10-Feb-2018 UK33S12057071-1-2

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1121939	
Customer Reference Number:	706514	
Product Description:	1000ML T/S BOTTLE DV5171 25-45kGy	
Validation Reference:	5171	
Quantity Received:	12	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	Bottle 009, B/N: 31012018, 1 plt	
	rradiation Data	
Date and Time of Irradiation:	10-Feb-2018 10:42	
Reference Dose Range kGy:	33.3 - 33.4	
Calculated Minimum Dose kGy:	32.0	
Calculated Maximum Dose kGy:	36.4	

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company



http://www.steris-ast.com

Certificate of Irradiation

Date issued: 09-Mar-2018 UK33S12069394-3-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1121921	
Customer Reference Number:	706678	
Product Description:	TRIGGER SPRAY HEAD CRYSTAL-G DV5170	
	25-45kGy	
Validation Reference:	5170	
Quantity Received:	20	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	45.0	
Customer Unit Lot/Batch Number:	20180221, 1 plt	
- I	rradiation Data	
Date and Time of Irradiation:	09-Mar-2018 14:45	
Reference Dose Range kGy:	28.7 - 34.9	
Calculated Minimum Dose kGy:	26.1	
Calculated Maximum Dose kGy:	38.6	

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company



http://www.steris-ast.com

Certificate of Irradiation

Date issued: 07-Feb-2018

UK33S12051056-1-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information		
Account Number:	100708	
Synergy Health Sales Part Reference:	1121938	
Customer Reference Number:	706435	
Product Description:	BAG 001-G DV5194 25-48kGy	
Validation Reference:	5194	
Quantity Received:	5	
Customer Minimum Specification kGy:	25.0	
Customer Maximum Specification kGy:	48.0	
Customer Unit Lot/Batch Number:	20171017, 1 PLT	
lı	rradiation Data	
Date and Time of Irradiation:	06-Feb-2018 18:57	
Reference Dose Range kGy:	38.4 - 38.4	
Calculated Minimum Dose kGy:	27.5	
Calculated Maximum Dose kGy:	42.8	

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company