



Certificate of Analysis-CRY 112

PRODUCT batch number: P808001

BASE batch number: P807800

ACTIVATOR batch number: P808000

Product Name:		Kryptonite Working Solution
Product batch number:		P808001
Reference:		CRY/SPO/021
This is to certify that Kryptonite Working Solution bearing the batch no's: P807800 and P808000 were tested on 2018-03-15 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear pale yellow liquid free of matter	Conforms
SPECIFIC GRAVITY	0.995 – 1.005	1.000
pH	2.5 – 5.5	3.0
ClO ₂ Concentration – after 10 minutes (ppm)	230 – 280	236 ppm

Product Name:		Kryptonite Base Solution
Reference:		CRY/SPO/020
Date of Manufacture:		2018-03-14
Date of Expiry:		2020-03-14
This is to certify that Kryptonite Base Solution was tested on 2018-03-15 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear blue liquid free of matter	Conforms
SPECIFIC GRAVITY	1.020 – 1.060	1.024
pH	1.5 – 3.5	1.8
STERILITY	No growth	Conforms





Product Name:	Kryptonite Activator Solution	
Reference:	CRY/SPO/019	
Date of Manufacture:	2018-03-14	
Date of Expiry:	2020-03-14	
This is to certify that Kryptonite Activator Solution was tested on 2018-03-15 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear and free of matter	Conforms
SPECIFIC GRAVITY	1.015 – 1.035	1.017
pH	11.5 – 13.0	11.8
SODIUM CHLORITE	2.10% +/- 0.1%	2.15%
STERILITY	No growth	Conforms

Completed By:

Reviewed By:

Quality Department: E Andrews

Stielinko

Date: 2018-04-20

Date: 2018-04-20



Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY112 Kryptonite Base
Lot No: P807800
ICN: 0318025D
Subcontractor's Report Number: 2018054187

Lab Book Reference: ST03-154
Date Received: 22 Mar 2018
Date Completed: 17 Apr 2018
No. of samples: 10

Sterility Test – Membrane Filtration

Medium	Result
Tryptone Soy Broth	No growth
Fluid Thioglycollate Medium	No growth

Conclusions:


The samples submitted comply with the test for sterility in accordance with the client's methodology.

Results Reported by:


.....
Laboratory Technician/Designee: Daniel Jennings


Date: 17 APR 2018

Reviewed by:

- 
.....
 Quality Assurance Manager
 UK General Manager
 Technical Manager
 Laboratory Manager

Date: 17 APR 2018

QA Approved by:

- 
.....
 Quality Assurance
 UK General Manager
 Technical Manager

Date: 17 APR 2018

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	D. Jennings	17 Apr 2018

CONFIDENTIAL
ST03-154

ICN: 0318025D

Page 1 of 1
TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 17 Apr 2018

UK
Deacon Park
Moorgate Road
Knowsley
Liverpool L33 7RX
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA
124 Bernard E. Saint Jean Drive
East Falmouth
MA 02536-4445
USA
Tel: 001 508 540 3444
Fax: 001 508 540 8680

Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Mörfelden-Walldorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15

Web: www.acciuk.co.uk

Registered in England and Wales. Company Registration Number: BR002906

Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY112 Kryptonite Activator
Lot No: P808000
ICN: 0318025D
Subcontractor's Report Number: 2018054188

Lab Book Reference: ST03-153
Date Received: 22 Mar 2018
Date Completed: 17 Apr 2018
No. of samples: 10

Sterility Test - Membrane Filtration

Medium	Result
Tryptone Soy Broth	No growth
Fluid Thioglycollate Medium	No growth

Conclusions:


The samples submitted comply with the test for sterility in accordance with the client's methodology.

Results Reported by:


.....
Laboratory Technician/Designee: Daniel Jennings

Date: 17 APR 2018

Reviewed by:

- 
.....
 Quality Assurance Manager
 UK General Manager
 Technical Manager
 Laboratory Manager

Date: 17 APR 2018

QA Approved by:

- 
.....
 Quality Assurance
 UK General Manager
 Technical Manager

Date: 17 APR 2018

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	D. Jennings	17 Apr 2018

CONFIDENTIAL
ST03-153

ICN: 0318025D

Page 1 of 1
TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 17 Apr 2018

UK
Deacon Park
Mourgate Road
Knowsley
Liverpool L33 7RX
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA
124 Bernard E. Saint Jean Drive
East Falmouth
MA 02536-4445
USA
Tel: 001 508 540 3444
Fax: 001 508 540 8680

Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Miefelden-Walkdorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15