



Certificate of Analysis

Product Name:	Crystel Platinum CRY109	
Reference:	CRY/PLA/019	
Batch Number:	P810202	
Date of Manufacture:	2018-01-15	
Expiry Date:	2020-01-15	
This is to certify that Crystel Platinum solution above was tested on 2018-01-16 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear solution free of matter	Conforms
SPECIFIC GRAVITY	0.900 – 1.100	0.988
pH (neat)	9.0 – 10.0	9.9
STERILITY	No growth	Conforms

Completed By:

Reviewed By:

Quality Department: E. Andrews

[Signature]

Date: 2018-05-15

Date: 2018-05-15

DRF1399 CRY/PLA/007/Issue3/June 2012





Specialists in Endotoxin and Glucan Detection

Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY109 Platinum
Lot No: P810202
ICN: 0418018D
Subcontractor's Report Number: 2018070291

Lab Book Reference: ST03-159
Date Received: 16 APR 2018
Date Completed: 08 MAY 2018
No. of samples: 10

Sterility Test - Membrane Filtration

Medium	Result
Tryptone Soy Broth	NO GROWTH
Fluid Thioglycollate Medium	NO GROWTH

Conclusions:

The samples submitted comply with the test for sterility in accordance with the client's methodology.

Results Reported by:


.....
Laboratory Technician/Designee: L. Ellis

Date: 09 MAY 2018

Reviewed by:


.....
 Quality Assurance Manager
 UK General Manager
 Technical Manager
 Laboratory Manager

Date: 09 MAY 2018

QA Approved by:


.....
 Quality Assurance
 UK General Manager
 Technical Manager

Date: 09 MAY 2018

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
NA	1	Revision 1, no changes	L Ellis	08 MAY 2018

CONFIDENTIAL
ST03-159

ICN: 0418018D

Page 1 of 1
TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 08 MAY 2018

UK
Deacon Park
Moorgate Road
Knossley
Liverpool L33 7RX
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA
124 Bernard E. Saint Jean Drive
East Falmouth
MA 02536-4445
USA
Tel: 001 508 540 3444
Fax: 001 508 540 8680

Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Miefelden-Walldorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15

Web: www.accuik.co.uk

Registered in England and Wales. Company Registration Number: BR002906



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 25-Feb-2018

UK33S12061299-2-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information

Account Number:	100708
Synergy Health Sales Part Reference:	1121930
Customer Reference Number:	706585
Product Description:	100ML BOTTLE CAP DV5165 25-45kGy
Validation Reference:	5165
Quantity Received:	6
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	07022018, 1 PLT

Irradiation Data

Date and Time of Irradiation:	24-Feb-2018 19:32
Reference Dose Range kGy:	33.5 - 33.5
Calculated Minimum Dose kGy:	30.7
Calculated Maximum Dose kGy:	37.3

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 11-Mar-2018

UK33S12069394-2-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information

Account Number:	100708
Synergy Health Sales Part Reference:	1121497
Customer Reference Number:	706678
Product Description:	BOTTLE 100ML DV5155 25-45kGy
Validation Reference:	5155
Quantity Received:	26
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	22022018, 2 plts

Irradiation Data

Date and Time of Irradiation:	10-Mar-2018 23:02
Reference Dose Range kGy:	33.2 - 33.8
Calculated Minimum Dose kGy:	31.2
Calculated Maximum Dose kGy:	36.7

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 22-Feb-2018

UK33S12057071-3-1

This is to certify that AST Daventry Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information

Account Number:	100708
Synergy Health Sales Part Reference:	1121936
Customer Reference Number:	706514
Product Description:	BAG 004-G DV5175 25-45kGy
Validation Reference:	5175
Quantity Received:	6
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	Bag 004-G, B/N: 20180131, EXP: 31.01.20, 1 pit

Irradiation Data

Date and Time of Irradiation:	22-Feb-2018 13:44
Reference Dose Range kGy:	34.7 - 34.7
Calculated Minimum Dose kGy:	26.6
Calculated Maximum Dose kGy:	37.2

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company

Processing Site: Brunel Close, Drayton Fields Industrial Estate, Daventry, NN11 8RB Phone No: +44 (0) 1327 706111

Registered Office: Ground Floor Stella, Windmill Hill Business Park, Whitehill Way, Swindon, Wiltshire SN5 6NX, UNITED KINGDOM
Company Registered in England and Wales No: 01771333 VAT Number: GB 813038069