




Certificate of Analysis-CRY 112

PRODUCT batch number: P818401
BASE batch number: P818400
ACTIVATOR batch number: P818301

Product Name:	Kryptonite Working Solution	
Product batch number:	P818401	
Reference:	CRY/SPO/021	
This is to certify that Kryptonite Working Solution bearing the batch no's: P818400 and P818301 were tested on 2018-04-25 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear pale yellow liquid free of matter	Conforms
SPECIFIC GRAVITY	0.995 – 1.005	1.000
pH	2.5 – 5.5	3.0
ClO ₂ Concentration – after 10 minutes (ppm)	230 – 280	236 ppm

Product Name:	Kryptonite Base Solution	
Reference:	CRY/SPO/020	
Date of Manufacture:	2018-04-25	
Date of Expiry:	2020-04-25	
This is to certify that Kryptonite Base Solution was tested on 2018-04-25 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear blue liquid free of matter	Conforms
SPECIFIC GRAVITY	1.020 – 1.060	1.020
pH	1.5 – 3.5	1.9
STERILITY	No growth	Conforms





Product Name:	Kryptonite Activator Solution	
Reference:	CRY/SPO/019	
Date of Manufacture:	2018-04-25	
Date of Expiry:	2020-04-25	
This is to certify that Kryptonite Activator Solution was tested on 2018-04-25 and found to comply with all requirements as set out in the specifications and method of analysis.		
PROPERTIES	SPECIFICATION	RESULTS
APPEARANCE	Clear and free of matter	Conforms
SPECIFIC GRAVITY	1.015 – 1.035	1.015
pH	11.5 – 13.0	12.0
SODIUM CHLORITE	2.10% +/- 0.1%	2.10%
STERILITY	No growth	Conforms

Completed By:


Reviewed By:

Quality Department: Carolews

[Signature]

Date: 2018-08-01

Date: 2018-08-01





ASSOCIATES OF
CAPE COD
INTERNATIONAL

Specialists in Endotoxin and Glucan Detection

Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY112 Kryptonite Base
Lot No: P818400
ICN: 0718006D
Subcontractor's Report Number: 2018124044

Lab Book Reference: ST03-177
Date Received: 05 Jul 2018
Date Completed: 25 Jul 2018
No. of samples: 10

Sterility Test – Membrane Filtration

Medium	Result
Tryptone Soy Broth	NO GROWTH
Fluid Thioglycollate Medium	NO GROWTH

Conclusions:


The samples submitted comply with the test for sterility in accordance with the client's methodology.

Results Reported by:


.....
Laboratory Technician/Designee: C.Thornton


Date: 25 Jul 2018

Reviewed by:


.....
 Quality Assurance Manager
 UK General Manager
 Technical Manager
 Laboratory Manager

Date: 25 Jul 2018

QA Approved by:


.....
 Quality Assurance
 UK General Manager
 Technical Manager

Date: 25 Jul 2018

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	C.Thornton	25 JUL 2018

CONFIDENTIAL
ST03-177

ICN: 0718006D

Page 1 of 1
TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 25 Jul 2018

UK
Deacon Park
Moongate Road
Knowsley
Liverpool L33 7RX
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA
124 Bernard E. Saint Jean Drive
East Falmouth
MA 02536-4445
USA
Tel: 001 508 540 3444
Fax: 001 508 540 8680

Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Morfelden-Walldorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15

Web: www.acciuk.co.uk

Registered in England and Wales. Company Registration Number: BR002906

Sterility Testing Analytical Report

Please note the results below are not generated at Associates of Cape Cod Int. Inc., but reported from a subcontractor's final report.

Sample ID: CRY112 Kryptonite Activator
Lot No: P818301
ICN: 0718006D
Subcontractor's Report Number: 2018124043

Lab Book Reference: ST03-178
Date Received: 05 Jul 2018
Date Completed: 25 Jul 2018
No. of samples: 10

Sterility Test - Membrane Filtration

Medium	Result
Tryptone Soy Broth	NO GROWTH
Fluid Thioglycollate Medium	NO GROWTH

Conclusions:

The samples submitted comply with the test for sterility in accordance with the client's methodology.

Results Reported by:

.....
Laboratory Technician/Designee: C.Thornton

Date: 25 Jul 2018

Reviewed by:

-
 Quality Assurance Manager
 UK General Manager
 Technical Manager
 Laboratory Manager

Date: 25 Jul 2018

QA Approved by:

-
 Quality Assurance
 UK General Manager
 Technical Manager

Date: 25 Jul 2018

Revision History:

DCCF No:	Revision:	Description:	Originator:	Effective Date:
N/A	1	Revision 1, no changes	C.Thornton	25 JUL 2018

CONFIDENTIAL
Error! No text of specified style in document.

ICN: 0718006D

Page 1 of 1
TSCTS044-1 Rev 3

REVISION NO: 1

DATE: 25 Jul 2018

UK
Deacon Park
Moorgate Road
Knowsley
Liverpool L33 7RX
Tel: +44 (0)151 547 7444
Fax: +44 (0)151 547 7400

USA
124 Bernard E. Saint Jean Drive
East Falmouth
MA 02536-4445
USA
Tel: 001 508 540 3444
Fax: 001 508 540 8680

Germany
PYROQUANT DIAGNOSTIK GmbH
Opelstrasse 14
D-64546 Mörfelden-Walldorf
Germany
Tel: 0049 6105 96 10 0
Fax: 0049 6105 96 10 15