







Certificate of Analysis-TITANIUM WIPES-CRY/TTM/049

PRODUCT batch number: P819700

Product Name:		TITANIUM WIPES Working Solution			
Reference:		CRY/TTM/047			
This is to certify that TITANIUM WIPES Working Solution bearing the Batch No's: P819700 were tested on 2018-07-05 and found to comply with all requirements set out in the specifications and method of analysis.					
PROPERTIES	SPECIFICATION		RESULTS		
APPEARANCE	Clear pale yellow liquid free of matter				Conforms
ODOUR	Chlorine like		Conforms		
ClO ₂ Concentration (ppm)	Conforms to QCF CRY/TTM/050		Conforms		

Product Name:		TITANIUM WIPES Base Solution		
Reference:		CRY/TTM/045		
Date of Manufacture:		2018-07-05		
Date of Expiry:		2020-07-05		
This is to certify that TITANIUM Wipes Base Solution 818600 was tested on 2018- 07-05 and found to comply with all requirements as set out in the specifications and method of analysis.				
PROPERTIES	SPECIFICATION		RESULTS	
APPEARANCE	Clear liquid free of matter		Conforms	
SPECIFIC GRAVITY	1.000-1.010		1.006	
рН	2.0 – 3.0		2.2	
STERILITY	No growth		Conforms	

DRF2550 CRY/TTM/049/Issue1











Product Name:		TITANIUM WIPES Activator Solution		
Reference:		CRY/TTM/046		
Date of Manufacture:		2018-05-04		
Date of Expiry:		2020-05-04		
This is to certify that TITA 2018-05-04 and found to specifications and metho	comply with all			
PROPERTIES		ICATION	RESULTS	
APPEARANCE	Clear and fr	ee of matter	Conforms	
рН	11.0	-11.5	11.2	
SODIUM CHLORITE	0.2% +/	- 0.005%	0.198%	
STERILITY	Nog	rowth	Conforms	

507		55 55	077.2250
Cai	mnl	etec	Rv.

Quality Department: Quality Department:

Date: 2018 - 09 - 03

Reviewed By:

Date: 2018-09-03

DRF2550 CRY/TTM/049/Issue1







Tristel Solutions Ltd Unit 1B, Lynx Business Park Fordham Road Cambridgeshire CB8 7NY

Certificate of Analysis

Sample Details

Product Name:

Titanium burstable wipes

Batch Number: Batch Code P819700

P819700

Job Number:

23508

Sample ID:

82240

Registered date:

10-Aug-18

Completion date:

28-Aug-18

Data

Test Method	Test	Specification	Result	Units	Pass/Fail
Direct Inoculation Sterility Test (cFTM)	DI Sterility (cFTM)	Complies / Does Not Cornply	Complies	N/A	Pass
Direct Inoculation Sterility Test (TSB)	DI Sterility (TSB)	Complies / Does Not Cornply	Complies	N/A	Pass
Membrane Filtration Sterility Test (cFTM)	MF Sterility (cFTM)	Complies / Does Not Cornply	Complies	N/A	Pass
Membrane Filtration Sterility Test (TSB)	MF Sterility (TSB)	Complies / Does Not Cornply	Complies	N/A	Pass

The product meets the following specification:

Sterility Test current Ph. Eur. by Membrane Filtration and Direct Inoculation. Validation report HVAL_R405.v2

LabHQ Product Specification Version:

.

Approval

Approved By

Elizabeth Parry

Approved Date:

28-Aug-2018 13:28:04

User ID

ераггу

Title

Quality Manager

This approval is performed by electronic signature in LabHQ LIMS and is the legal equivalent of the approver's hand-written signature.



http://www.steris-ast.com

Certificate of Irradiation

Date Issued: 26-Jul-2018

UK31S12142101-1-1

This is to certify that Elgin Synergy Health PLC, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products EN ISO 13485 Quality System - Medical Devices

The Tristel Company Ltd
Unit 4C Lynx Bus Park, Fordham Rd
Newmarket
Suffolk CB8 7NY
UNITED KINGDOM

Order Information			
Account Number: Synergy Health Sales Part Reference: Customer Reference Number: Product Description: Validation Reference: Quantity Received: Customer Minimum Specification kGy: Customer Maximum Specification kGy: Customer Unit Lot/Batch Number:	100708 1103783 707714 BURST/TITANIUM WIPES 10-20KGY 1.6922 100 10.0 20.0 CRY144, LOT P819700, 3 PALLETS		
li li	rradiation Data		
Date and Time of Irradiation: Reference Dose Range kGy: Calculated Minimum Dose kGy: Calculated Maximum Dose kGy:	26-Jul-2018 22:03 17.7 - 18.9 14.8 19.0		

Irradiation Release Authorised By Synergy Health Sterilisation UK, a STERIS Company